



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
Data File : VX038828.D
Acq On : 16 Nov 2023 08:31
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Title : SW846 8260
Last Update : Thu Nov 16 22:59:22 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	9528	PASS
75	95	30	60	48.9	26217	PASS
95	95	100	100	100.0	53657	PASS
96	95	5	9	6.1	3284	PASS
173	174	0.00	2	0.8	262	PASS
174	95	50	100	64.6	34656	PASS
175	174	5	9	7.3	2516	PASS
176	174	95	101	97.2	33696	PASS
177	176	5	9	7.0	2351	PASS