



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
Data File : VX038840.D
Acq On : 16 Nov 2023 15:59
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Title : SW846 8260
Last Update : Thu Nov 16 22:59:22 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	9933	PASS
75	95	30	60	51.1	27027	PASS
95	95	100	100	100.0	52914	PASS
96	95	5	9	7.1	3783	PASS
173	174	0.00	2	0.7	259	PASS
174	95	50	100	65.5	34659	PASS
175	174	5	9	7.3	2527	PASS
176	174	95	101	95.8	33219	PASS
177	176	5	9	6.6	2198	PASS