

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111718\
 Data File : VX006071.D
 Acq On : 17 Nov 2018 16:25
 Operator : JC/MD
 Sample : J5934-09DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-0145DL

Quant Time: Nov 19 09:14:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	124485	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	198692	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	152448	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	95634	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	76106	53.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.36%	
35) Dibromofluoromethane	5.50	113	61015	45.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.30%	
50) Toluene-d8	8.71	98	200198	44.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.12%	
62) 4-Bromofluorobenzene	11.14	95	69256	42.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.74%	
Target Compounds						
40) Benzene	6.14	78	30950	5.54	ug/l	Qvalue # 91
65) Chlorobenzene	10.14	112	220883	72.33	ug/l	99
67) Ethyl Benzene	10.25	91	16034	3.27	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111718\
Data File : VX006071.D
Acq On : 17 Nov 2018 16:25
Operator : JC/MD
Sample : J5934-09DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-0145DL

Quant Time: Nov 19 09:14:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

