

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111722\
 Data File : VX032896.D
 Acq On : 17 Nov 2022 13:44
 Operator : JC/MD
 Sample : N5671-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D1-20221115

Quant Time: Nov 18 01:40:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	181186	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76897	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77171	50.503	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.000%		
35) Dibromofluoromethane	5.385	113	63158	48.868	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.740%		
50) Toluene-d8	8.647	98	225667	50.215	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.420%		
62) 4-Bromofluorobenzene	11.079	95	88172	50.227	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 100.460%		
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	3.117	73	1417	0.366	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	18396	13.257	ug/l	87
44) Trichloroethene	7.129	130	162629	113.432	ug/l	100
52) Toluene	8.720	92	1808	0.544	ug/l	100
64) Tetrachloroethene	9.275	164	17678	14.699	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111722\
Data File : VX032896.D
Acq On : 17 Nov 2022 13:44
Operator : JC/MD
Sample : N5671-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D1-20221115

Quant Time: Nov 18 01:40:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

