

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111722\
 Data File : VX032899.D
 Acq On : 17 Nov 2022 14:53
 Operator : JC/MD
 Sample : N5671-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178I-20221115

Quant Time: Nov 18 01:41:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178356	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77330	51.095	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.200%		
35) Dibromofluoromethane	5.385	113	61340	48.215	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.420%		
50) Toluene-d8	8.647	98	219885	49.705	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.400%		
62) 4-Bromofluorobenzene	11.079	95	84925	49.145	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.280%		
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.111	73	2653	0.691	ug/l #	82
20) Methylene Chloride	2.788	84	391	0.310	ug/l #	43
30) Chloroform	5.105	83	1137	0.478	ug/l	87
44) Trichloroethene	7.129	130	2862	2.028	ug/l	99
52) Toluene	8.720	92	941	0.288	ug/l	96
64) Tetrachloroethene	9.269	164	737	0.633	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111722\
Data File : VX032899.D
Acq On : 17 Nov 2022 14:53
Operator : JC/MD
Sample : N5671-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178I-20221115

Quant Time: Nov 18 01:41:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

