

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111722\
 Data File : VX032910.D
 Acq On : 17 Nov 2022 19:07
 Operator : JC/MD
 Sample : N5672-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20221116

Quant Time: Nov 18 01:42:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	184498	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81991	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79069	51.126	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.260%		
35) Dibromofluoromethane	5.385	113	63316	48.111	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.220%		
50) Toluene-d8	8.647	98	234588	51.263	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.520%		
62) 4-Bromofluorobenzene	11.079	95	90081	50.393	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 100.780%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	11006	9.211	ug/l	99
12) 1,1-Dichloroethene	2.319	96	1512	1.355	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	2808	1.999	ug/l	90
30) Chloroform	5.093	83	1041	0.429	ug/l	98
44) Trichloroethene	7.129	130	928039	635.676	ug/l	99
52) Toluene	8.720	92	1741	0.515	ug/l	93
55) 1,1,2-Trichloroethane	9.153	97	457	0.321	ug/l #	78
64) Tetrachloroethene	9.275	164	7726	6.256	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111722\
Data File : VX032910.D
Acq On : 17 Nov 2022 19:07
Operator : JC/MD
Sample : N5672-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20221116

Quant Time: Nov 18 01:42:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

