

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038914.D
 Acq On : 18 Nov 2023 07:03
 Operator : JC/MD
 Sample : 05404-13MS
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20231109MS

Quant Time: Nov 20 00:34:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	126151	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	242790	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103964	51.145	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.300%
35) Dibromofluoromethane	5.385	113	84024	47.918	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.840%
50) Toluene-d8	8.646	98	313614	46.746	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.500%
62) 4-Bromofluorobenzene	11.079	95	124329	46.294	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.580%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	76215	51.778	ug/l	100
3) Chloromethane	1.294	50	88041	45.826	ug/l	98
4) Vinyl Chloride	1.373	62	90300	48.794	ug/l	99
5) Bromomethane	1.599	94	43449	56.784	ug/l	100
6) Chloroethane	1.678	64	54635	51.732	ug/l	100
7) Trichlorofluoromethane	1.885	101	122532	52.177	ug/l	99
8) Diethyl Ether	2.135	74	50403	49.278	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.324	101	72585	46.607	ug/l	97
10) Methyl Iodide	2.452	142	87674	46.253	ug/l	93
11) Tert butyl alcohol	2.965	59	161374	285.433	ug/l #	92
12) 1,1-Dichloroethene	2.318	96	73473	46.729	ug/l	89
13) Acrolein	2.233	56	80412	268.834	ug/l	100
14) Allyl chloride	2.660	41	146337	46.486	ug/l	93
15) Acrylonitrile	3.062	53	308998	258.659	ug/l	99
16) Acetone	2.379	43	265619	245.466	ug/l	91
17) Carbon Disulfide	2.507	76	202849	41.649	ug/l	98
18) Methyl Acetate	2.702	43	237010	46.112	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	280585	50.737	ug/l	97
20) Methylene Chloride	2.788	84	88587	45.414	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	80957	46.664	ug/l	95
22) Diisopropyl ether	3.763	45	288366	49.933	ug/l #	82
23) Vinyl Acetate	3.721	43	900001	216.386	ug/l	98
24) 1,1-Dichloroethane	3.605	63	159168	49.286	ug/l	99
25) 2-Butanone	4.556	43	449346	256.071	ug/l	94
26) 2,2-Dichloropropane	4.476	77	90967	32.473	ug/l	94
27) cis-1,2-Dichloroethene	4.489	96	97482	48.205	ug/l	93
28) Bromochloromethane	4.897	49	79377	51.193	ug/l	100
29) Tetrahydrofuran	5.001	42	296612	266.215	ug/l	97
30) Chloroform	5.092	83	161128	51.095	ug/l	95
31) Cyclohexane	5.470	56	136848	47.153	ug/l	98
32) 1,1,1-Trichloroethane	5.385	97	138610	50.746	ug/l	99
36) 1,1-Dichloropropene	5.690	75	115475	45.948	ug/l	98
37) Ethyl Acetate	4.714	43	143604	43.859	ug/l	98
38) Carbon Tetrachloride	5.677	117	116641	50.217	ug/l	97
39) Methylcyclohexane	7.378	83	139466	42.544	ug/l	96
40) Benzene	6.037	78	347418	46.739	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038914.D
 Acq On : 18 Nov 2023 07:03
 Operator : JC/MD
 Sample : 05404-13MS
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20231109MS

Quant Time: Nov 20 00:34:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	84815	49.945	ug/l	93
42) 1,2-Dichloroethane	6.086	62	124708	50.638	ug/l	97
43) Isopropyl Acetate	6.342	43	239881	45.842	ug/l	98
44) Trichloroethene	7.122	130	92682	51.226	ug/l	95
45) 1,2-Dichloropropane	7.427	63	95413	47.699	ug/l	97
46) Dibromomethane	7.580	93	66157	49.594	ug/l	99
47) Bromodichloromethane	7.823	83	127385	48.113	ug/l	95
48) Methyl methacrylate	7.689	41	121157	47.721	ug/l	93
49) 1,4-Dioxane	7.659	88	57758	1175.909	ug/l #	95
51) 4-Methyl-2-Pentanone	8.573	43	850104	254.217	ug/l	99
52) Toluene	8.720	92	218211	47.307	ug/l	96
53) t-1,3-Dichloropropene	8.976	75	140328	44.600	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	145436	44.750	ug/l	91
55) 1,1,2-Trichloroethane	9.152	97	92310	47.962	ug/l	93
56) Ethyl methacrylate	9.116	69	159302	48.919	ug/l	95
57) 1,3-Dichloropropane	9.305	76	154641	48.582	ug/l	100
59) 2-Hexanone	9.427	43	686999	256.313	ug/l	98
60) Dibromochloromethane	9.518	129	97185	49.661	ug/l	99
61) 1,2-Dibromoethane	9.610	107	98702	48.872	ug/l	99
64) Tetrachloroethene	9.274	164	63831	45.038	ug/l	97
65) Chlorobenzene	10.079	112	229477	46.721	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.164	131	85334	48.299	ug/l	98
67) Ethyl Benzene	10.195	91	425631	47.167	ug/l	100
68) m/p-Xylenes	10.299	106	323997	94.915	ug/l	100
69) o-Xylene	10.640	106	160479	47.512	ug/l	100
70) Styrene	10.652	104	266694	46.753	ug/l	98
71) Bromoform	10.799	173	72295	49.084	ug/l #	100
73) Isopropylbenzene	10.963	105	415922	46.384	ug/l	100
74) N-amyl acetate	10.841	43	195205	41.182	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	162895	46.537	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	183014	63.737	ug/l	89
77) Bromobenzene	11.195	156	93468	46.390	ug/l	97
78) n-propylbenzene	11.305	91	503049	45.577	ug/l	99
79) 2-Chlorotoluene	11.365	91	294785	45.004	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	352564	46.762	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	49229	38.028	ug/l #	83
82) 4-Chlorotoluene	11.451	91	346202	45.838	ug/l	100
83) tert-Butylbenzene	11.713	119	353989	47.673	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	349170	45.756	ug/l	99
85) sec-Butylbenzene	11.890	105	441637	45.470	ug/l	99
86) p-Isopropyltoluene	12.006	119	355234	45.561	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	180367	45.714	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	183530	45.448	ug/l	99
89) n-Butylbenzene	12.329	91	337858	42.949	ug/l	98
90) Hexachloroethane	12.536	117	71669	46.030	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	178079	47.586	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.944	75	41774	48.710	ug/l	91
93) 1,2,4-Trichlorobenzene	13.585	180	110558	44.087	ug/l	97
94) Hexachlorobutadiene	13.725	225	36593	41.206	ug/l	98
95) Naphthalene	13.774	128	429832	46.022	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	110418	46.449	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038914.D
Acq On : 18 Nov 2023 07:03
Operator : JC/MD
Sample : 05404-13MS
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Quant Time: Nov 20 00:34:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#)=qualifier out of range (m)=manual integration (+)=signals summed						


Quant Time: Nov 20 00:34:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.549 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS







Tgt Ion:168 Resp: 126151
Ion Ratio Lower Upper
168 100
99 61.9 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 51.778 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 85 Resp: 76215
Ion Ratio Lower Upper
85 100
87 33.6 16.7 50.1









#3
Chloromethane
Concen: 45.826 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 50 Resp: 88041
Ion Ratio Lower Upper
50 100
52 32.8 25.5 38.3



#4
Vinyl Chloride
Concen: 48.794 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.001 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 62 Resp: 90300
Ion Ratio Lower Upper
62 100
64 31.0 25.4 38.0









#5
Bromomethane
Concen: 56.784 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 94 Resp: 43449
Ion Ratio Lower Upper
94 100
96 93.6 74.6 111.8



#6
Chloroethane
Concen: 51.732 ug/l
RT: 1.678 min Scan# 97
Delta R.T. -0.007 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 64 Resp: 54635
Ion Ratio Lower Upper
64 100
66 31.5 25.1 37.7









#7
 Trichlorofluoromethane
 Concen: 52.177 ug/l
 RT: 1.885 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX038914.D
 Acq: 18 Nov 2023 07:03




Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20231109MS


Tgt Ion:101 Resp: 122532
 Ion Ratio Lower Upper
 101 100
 103 63.6 51.4 77.2



#8
 Diethyl Ether
 Concen: 49.278 ug/l
 RT: 2.135 min Scan# 172
 Delta R.T. 0.006 min
 Lab File: VX038914.D
 Acq: 18 Nov 2023 07:03

Tgt Ion: 74 Resp: 50403
 Ion Ratio Lower Upper
 74 100
 45 102.8 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 46.607 ug/l
RT: 2.324 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion:101 Resp: 72585
Ion Ratio Lower Upper
101 100
85 46.4 37.0 55.6
151 69.2 58.6 87.8



#10
Methyl Iodide
Concen: 46.253 ug/l
RT: 2.452 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion:142 Resp: 87674
Ion Ratio Lower Upper
142 100
127 45.5 41.6 62.4
141 14.2 11.6 17.4





#11

Tert butyl alcohol

Concen: 285.433 ug/l

RT: 2.965 min Scan# 306

Delta R.T. 0.012 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :





RE134D1-20231109MS

Tgt Ion: 59 Resp: 161374

Ion Ratio Lower Upper

59 100

57 10.4 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 46.729 ug/l

RT: 2.318 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion: 96 Resp: 73473


Ion Ratio Lower Upper

96 100

61 160.3 144.6 216.8

98 64.8 50.4 75.6









#13
Acrolein
Concen: 268.834 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 56 Resp: 80412
Ion Ratio Lower Upper
56 100
55 71.2 57.1 85.7



#14
Allyl chloride
Concen: 46.486 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 41 Resp: 146337
Ion Ratio Lower Upper
41 100
39 61.9 55.8 83.6
76 31.9 24.3 36.5









#15
Acrylonitrile
Concen: 258.659 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 53 Resp: 308998
Ion Ratio Lower Upper
53 100
52 82.4 66.0 99.0
51 35.2 29.8 44.6



#16
Acetone
Concen: 245.466 ug/l
RT: 2.379 min Scan# 212
Delta R.T. 0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 43 Resp: 265619
Ion Ratio Lower Upper
43 100
58 32.2 21.9 32.9





#17

Carbon Disulfide

Concen: 41.649 ug/l

RT: 2.507 min Scan# 21

Delta R.T. -0.006 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :





RE134D1-20231109MS

Tgt Ion: 76 Resp: 202849

Ion Ratio Lower Upper

76 100

78 9.5 7.0 10.6



#18

Methyl Acetate

Concen: 46.112 ug/l

RT: 2.702 min Scan# 265

Delta R.T. 0.006 min

Lab File: VX038914.D




Acq: 18 Nov 2023 07:03


Tgt Ion: 43 Resp: 237010

Ion Ratio Lower Upper

43 100

74 23.6 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 50.737 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 73 Resp: 280585
Ion Ratio Lower Upper
73 100
57 22.9 17.3 25.9



#20
Methylene Chloride
Concen: 45.414 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 84 Resp: 88587
Ion Ratio Lower Upper
84 100
49 126.4 103.6 155.4
51 38.3 33.8 50.6
86 62.4 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 46.664 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 96 Resp: 80957
Ion Ratio Lower Upper
96 100
61 142.0 119.2 178.8
98 64.1 48.7 73.1



#22
Diisopropyl ether
Concen: 49.933 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 45 Resp: 288366
Ion Ratio Lower Upper
45 100
43 64.5 69.7 104.5#
87 27.5 20.8 31.2
59 11.4 9.0 13.4









#23
 Vinyl Acetate
 Concen: 216.386 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX038914.D
 Acq: 18 Nov 2023 07:03




Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20231109MS


Tgt Ion: 43 Resp: 900001
 Ion Ratio Lower Upper
 43 100
 86 10.4 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 49.286 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX038914.D
 Acq: 18 Nov 2023 07:03

Tgt Ion: 63 Resp: 159168
 Ion Ratio Lower Upper
 63 100
 98 6.1 3.3 9.8
 100 4.3 2.1 6.5









#25
2-Butanone
Concen: 256.071 ug/l
RT: 4.556 min Scan# 50
Delta R.T. 0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 43 Resp: 449346
Ion Ratio Lower Upper
43 100
72 27.1 19.4 29.2



#26
2,2-Dichloropropane
Concen: 32.473 ug/l
RT: 4.476 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 77 Resp: 90967
Ion Ratio Lower Upper
77 100
97 24.4 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 48.205 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS





Tgt Ion: 96 Resp: 97482

Ion Ratio Lower Upper

96 100

61 142.7 0.0 312.2

98 64.6 0.0 129.0



#28

Bromochloromethane

Concen: 51.193 ug/l

RT: 4.897 min Scan# 625

Delta R.T. 0.006 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion: 49 Resp: 79377


Ion Ratio Lower Upper

49 100

129 0.9 0.0 3.8

130 71.7 57.5 86.3





#29

Tetrahydrofuran

Concen: 266.215 ug/l

RT: 5.001 min Scan# 641

Delta R.T. 0.006 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS





Tgt Ion: 42 Resp: 296612

Ion Ratio Lower Upper

42 100

72 44.7 34.2 51.4

71 41.6 32.2 48.2



#30

Chloroform

Concen: 51.095 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.000 min

Lab File: VX038914.D




Acq: 18 Nov 2023 07:03


Tgt Ion: 83 Resp: 161128

Ion Ratio Lower Upper

83 100


85 66.5 50.3 75.5








#31
Cyclohexane
Concen: 47.153 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03


Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS





Tgt Ion: 56 Resp: 136848
Ion Ratio Lower Upper
56 100
69 29.2 23.4 35.2
84 80.6 65.9 98.9




#32
1,1,1-Trichloroethane
Concen: 50.746 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03



Tgt Ion: 97 Resp: 138610
Ion Ratio Lower Upper
97 100
99 63.6 51.1 76.7
61 43.9 36.4 54.6









#33
 1,2-Dichloroethane-d4
 Concen: 51.145 ug/l
 RT: 5.958 min Scan# 799
 Delta R.T. -0.000 min
 Lab File: VX038914.D
 Acq: 18 Nov 2023 07:03




Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20231109MS


Tgt Ion: 65 Resp: 103964
 Ion Ratio Lower Upper
 65 100
 67 53.8 0.0 100.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.763 min Scan# 931
 Delta R.T. -0.000 min
 Lab File: VX038914.D
 Acq: 18 Nov 2023 07:03

Tgt Ion: 114 Resp: 242790
 Ion Ratio Lower Upper
 114 100
 63 20.9 0.0 45.0
 88 16.4 0.0 36.6









#35
Dibromofluoromethane
Concen: 47.918 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion:113 Resp: 84024
Ion Ratio Lower Upper
113 100
111 103.0 82.9 124.3
192 16.1 13.3 19.9



#36
1,1-Dichloropropene
Concen: 45.948 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 75 Resp: 115475
Ion Ratio Lower Upper
75 100
110 35.6 17.0 50.9
77 31.4 24.7 37.1





#37

Ethyl Acetate

Concen: 43.859 ug/l

RT: 4.714 min Scan# 594

Delta R.T. 0.006 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS





Tgt Ion: 43 Resp: 143604

Ion Ratio Lower Upper

43 100

61 14.5 10.6 16.0

70 10.9 8.3 12.5



#38

Carbon Tetrachloride

Concen: 50.217 ug/l

RT: 5.677 min Scan# 753

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion:117 Resp: 116641


Ion Ratio Lower Upper

117 100

119 94.2 78.2 117.2

121 31.8 24.4 36.6









#39
Methylcyclohexane
Concen: 42.544 ug/l
RT: 7.378 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 83 Resp: 139466
Ion Ratio Lower Upper
83 100
55 77.5 64.6 97.0
98 43.7 36.7 55.1



#40
Benzene
Concen: 46.739 ug/l
RT: 6.037 min Scan# 812
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 78 Resp: 347418
Ion Ratio Lower Upper
78 100
77 23.6 19.3 28.9









#41
Methacrylonitrile
Concen: 49.945 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion:	41	Resp:	84815
Ion	Ratio	Lower	Upper
41	100		
39	55.1	45.2	67.8
67	75.0	52.6	78.8
52	30.4	22.7	34.1



#42
1,2-Dichloroethane
Concen: 50.638 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion:	62	Resp:	124708
Ion	Ratio	Lower	Upper
62	100		
98	9.0	0.0	16.0





#43

Isopropyl Acetate

Concen: 45.842 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS





Tgt Ion: 43 Resp: 239881

Ion Ratio Lower Upper

43 100

61 19.9 15.2 22.8

87 12.9 9.8 14.6



#44

Trichloroethene

Concen: 51.226 ug/l

RT: 7.122 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX038914.D




Acq: 18 Nov 2023 07:03


Tgt Ion:130 Resp: 92682

Ion Ratio Lower Upper

130 100

95 107.4 0.0 204.4









#45
1,2-Dichloropropane
Concen: 47.699 ug/l
RT: 7.427 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 63 Resp: 95413
Ion Ratio Lower Upper
63 100
65 31.9 24.3 36.5



#46
Dibromomethane
Concen: 49.594 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 93 Resp: 66157
Ion Ratio Lower Upper
93 100
95 83.1 66.6 100.0
174 84.5 69.4 104.0









#47
Bromodichloromethane
Concen: 48.113 ug/l
RT: 7.823 min Scan# 1104
Delta R.T. 0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 83 Resp: 127385
Ion Ratio Lower Upper
83 100
85 66.1 49.8 74.6
127 8.1 7.1 10.7



#48
Methyl methacrylate
Concen: 47.721 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 41 Resp: 121157
Ion Ratio Lower Upper
41 100
69 83.6 59.2 88.8
39 58.4 47.7 71.5









#49
1,4-Dioxane
Concen: 1175.909 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 88 Resp: 57758
Ion Ratio Lower Upper
88 100
43 31.2 0.0 0.0#
58 70.8 60.1 90.1



#50
Toluene-d8
Concen: 46.746 ug/l
RT: 8.646 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 98 Resp: 313614
Ion Ratio Lower Upper
98 100
100 64.7 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 254.217 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 43 Resp: 850104
Ion Ratio Lower Upper
43 100
58 38.7 30.4 45.6



#52
Toluene
Concen: 47.307 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 92 Resp: 218211
Ion Ratio Lower Upper
92 100
91 172.9 134.2 201.2









#53
t-1,3-Dichloropropene
Concen: 44.600 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 75 Resp: 140328
Ion Ratio Lower Upper
75 100
77 32.0 25.7 38.5



#54
cis-1,3-Dichloropropene
Concen: 44.750 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03


Tgt Ion: 75 Resp: 145436
Ion Ratio Lower Upper
75 100
77 31.5 25.3 37.9
39 48.6 46.7 70.1








#55
1,1,2-Trichloroethane
Concen: 47.962 ug/l
RT: 9.152 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03


Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS





Tgt Ion:	97	Resp:	92310
Ion	Ratio	Lower	Upper
97	100		
83	89.6	65.8	98.8
85	57.1	41.6	62.4
99	64.7	47.8	71.8




#56
Ethyl methacrylate
Concen: 48.919 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03



Tgt Ion:	69	Resp:	159302
Ion	Ratio	Lower	Upper
69	100		
41	66.5	56.8	85.2
39	37.5	32.0	48.0





#57

1,3-Dichloropropane

Concen: 48.582 ug/l

RT: 9.305 min Scan# 1348

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :





RE134D1-20231109MS

Tgt Ion: 76 Resp: 154641

Ion Ratio Lower Upper

76 100

78 32.8 26.1 39.1



#59

2-Hexanone

Concen: 256.313 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX038914.D




Acq: 18 Nov 2023 07:03


Tgt Ion: 43 Resp: 686999

Ion Ratio Lower Upper

43 100

58 53.9 26.2 78.5









#60
Dibromochloromethane
Concen: 49.661 ug/l
RT: 9.518 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion:129 Resp: 97185
Ion Ratio Lower Upper
129 100
127 77.0 38.8 116.4



#61
1,2-Dibromoethane
Concen: 48.872 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion:107 Resp: 98702
Ion Ratio Lower Upper
107 100
109 93.5 75.2 112.8





#62

4-Bromofluorobenzene

Concen: 46.294 ug/l

RT: 11.079 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS





Tgt Ion: 95 Resp: 124329

Ion Ratio Lower Upper

95 100

174 69.0 0.0 135.6

176 66.0 0.0 131.0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion: 117 Resp: 228012


Ion Ratio Lower Upper

117 100

82 57.3 46.2 69.4

119 31.8 25.5 38.3





#64

Tetrachloroethene

Concen: 45.038 ug/l

RT: 9.274 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX038914.D


Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS



Tgt Ion:164 Resp: 63831




Ion Ratio Lower Upper

164 100

166 127.8 102.7 154.1

129 104.9 80.4 120.6

131 98.9 76.5 114.7



#65

Chlorobenzene

Concen: 46.721 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX038914.D




Acq: 18 Nov 2023 07:03


Tgt Ion:112 Resp: 229477

Ion Ratio Lower Upper

112 100

114 32.5 25.7 38.5









#66
1,1,1,2-Tetrachloroethane
Concen: 48.299 ug/l
RT: 10.164 min Scan# 1489
Delta R.T. 0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion:131 Resp: 85334
Ion Ratio Lower Upper
131 100
133 94.8 47.1 141.4
119 67.7 32.2 96.6



#67
Ethyl Benzene
Concen: 47.167 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 91 Resp: 425631
Ion Ratio Lower Upper
91 100
106 30.9 24.9 37.3









#68
m/p-Xylenes
Concen: 94.915 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion:106 Resp: 323997
Ion Ratio Lower Upper
106 100
91 206.5 165.5 248.3



#69
o-Xylene
Concen: 47.512 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion:106 Resp: 160479
Ion Ratio Lower Upper
106 100
91 217.5 108.9 326.8





#70

Styrene

Concen: 46.753 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS





Tgt Ion:104 Resp: 266694

Ion Ratio Lower Upper

104 100

78 52.7 43.6 65.4

103 55.5 43.6 65.4



#71

Bromoform

Concen: 49.084 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion:173 Resp: 72295


Ion Ratio Lower Upper

173 100

175 48.3 24.1 72.4

254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS





Tgt Ion:152 Resp: 115087

Ion Ratio Lower Upper

152 100

115 89.4 43.3 129.9

150 173.2 0.0 345.0



#73

Isopropylbenzene

Concen: 46.384 ug/l

RT: 10.963 min Scan# 1620

Delta R.T. -0.000 min

Lab File: VX038914.D




Acq: 18 Nov 2023 07:03


Tgt Ion:105 Resp: 415922

Ion Ratio Lower Upper

105 100

120 25.8 12.9 38.6





#74

N-amy1 acetate

Concen: 41.182 ug/l

RT: 10.841 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX038914.D


Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS



Tgt Ion: 43 Resp: 195205




Ion Ratio Lower Upper

43 100

70 40.3 32.9 49.3

55 22.6 19.3 28.9

61 23.8 18.8 28.2



#75

1,1,2,2-Tetrachloroethane

Concen: 46.537 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion: 83 Resp: 162895


Ion Ratio Lower Upper

83 100

131 9.1 5.3 15.8

85 63.8 32.0 96.2









#76
1,2,3-Trichloropropane
Concen: 63.737 ug/l
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 75 Resp: 183014
Ion Ratio Lower Upper
75 100
77 31.8 19.1 57.5



#77
Bromobenzene
Concen: 46.390 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 156 Resp: 93468
Ion Ratio Lower Upper
156 100
77 181.4 88.4 265.3
158 97.4 48.0 144.0









#78
n-propylbenzene
Concen: 45.577 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 91 Resp: 503049
Ion Ratio Lower Upper
91 100
120 22.2 11.5 34.4



#79
2-Chlorotoluene
Concen: 45.004 ug/l
RT: 11.365 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 91 Resp: 294785
Ion Ratio Lower Upper
91 100
126 32.4 16.3 48.8





#80

1,3,5-Trimethylbenzene

Concen: 46.762 ug/l

RT: 11.451 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :





RE134D1-20231109MS

Tgt Ion:105 Resp: 352564

Ion Ratio Lower Upper

105 100

120 46.0 24.1 72.3



#81

trans-1,4-Dichloro-2-butene

Concen: 38.028 ug/l

RT: 11.018 min Scan# 1629

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion: 75 Resp: 49229


Ion Ratio Lower Upper

75 100

53 87.7 90.5 135.7#

89 45.8 36.9 55.3





#82

4-Chlorotoluene

Concen: 45.838 ug/l

RT: 11.451 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :





RE134D1-20231109MS

Tgt Ion: 91 Resp: 346202

Ion Ratio Lower Upper

91 100

126 28.6 14.2 42.6



#83

tert-Butylbenzene

Concen: 47.673 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion: 119 Resp: 353989


Ion Ratio Lower Upper

119 100

91 60.3 31.2 93.6

134 21.1 11.7 35.0





#84

1,2,4-Trimethylbenzene

Concen: 45.756 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :





RE134D1-20231109MS

Tgt Ion:105 Resp: 349170

Ion Ratio Lower Upper

105 100

120 43.0 21.2 63.6



#85

sec-Butylbenzene

Concen: 45.470 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX038914.D




Acq: 18 Nov 2023 07:03


Tgt Ion:105 Resp: 441637

Ion Ratio Lower Upper

105 100

134 19.2 9.8 29.4





#86

p-Isopropyltoluene

Concen: 45.561 ug/l

RT: 12.006 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :

RE134D1-20231109MS





Tgt Ion:119 Resp: 355234

Ion Ratio Lower Upper

119 100

134 26.1 12.8 38.3

91 26.5 13.1 39.3



#87

1,3-Dichlorobenzene

Concen: 45.714 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion:146 Resp: 180367


Ion Ratio Lower Upper

146 100

111 44.4 22.1 66.1


148 64.4 31.8 95.3








#88
1,4-Dichlorobenzene
Concen: 45.448 ug/l
RT: 12.042 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS







Tgt Ion:146 Resp: 183530
Ion Ratio Lower Upper
146 100
111 44.1 21.3 63.7
148 63.9 31.8 95.4



#89
n-Butylbenzene
Concen: 42.949 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion: 91 Resp: 337858
Ion Ratio Lower Upper
91 100
92 53.0 27.4 82.0
134 24.1 12.6 37.8





#90

Hexachloroethane

Concen: 46.030 ug/l

RT: 12.536 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03

Instrument :

MSVOA_X

ClientSampleId :





RE134D1-20231109MS

Tgt Ion:117 Resp: 71669

Ion Ratio Lower Upper

117 100

201 63.2 32.9 98.6



#91

1,2-Dichlorobenzene

Concen: 47.586 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX038914.D

Acq: 18 Nov 2023 07:03




Tgt Ion:146 Resp: 178079


Ion Ratio Lower Upper

146 100

111 45.2 22.4 67.0

148 63.8 31.6 95.0









#92
1,2-Dibromo-3-Chloropropane
Concen: 48.710 ug/l
RT: 12.944 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion: 75 Resp: 41774
Ion Ratio Lower Upper
75 100
155 78.5 36.1 108.5
157 99.0 44.9 134.7



#93
1,2,4-Trichlorobenzene
Concen: 44.087 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03

Tgt Ion:180 Resp: 110558
Ion Ratio Lower Upper
180 100
182 93.9 48.3 144.8
145 33.3 18.0 54.0









#94
Hexachlorobutadiene
Concen: 41.206 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03




Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109MS


Tgt Ion:225 Resp: 36593
Ion Ratio Lower Upper
225 100
223 62.7 31.1 93.2
227 67.5 32.5 97.5



#95
Naphthalene
Concen: 46.022 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX038914.D
Acq: 18 Nov 2023 07:03


Tgt Ion:128 Resp: 429832
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.8 8.8 13.2





#96
 1,2,3-Trichlorobenzene
 Concen: 46.449 ug/l
 RT: 13.963 min Scan# 2112
 Delta R.T. 0.006 min
 Lab File: VX038914.D
 Acq: 18 Nov 2023 07:03

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20231109MS



Tgt Ion:180 Resp: 110418
 Ion Ratio Lower Upper
 180 100
 182 94.2 46.9 140.6
 145 35.5 17.8 53.5

