

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038875.D
 Acq On : 17 Nov 2023 13:10
 Operator : JC/MD
 Sample : 05403-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB02-20231107

Quant Time: Nov 17 22:49:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	254000	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	257095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94458	46.520	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.040%
35) Dibromofluoromethane	5.391	113	81347	44.344	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.680%
50) Toluene-d8	8.647	98	319319	45.496	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.000%#
62) 4-Bromofluorobenzene	11.079	95	135929	48.380	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.760%
Target Compounds						
16) Acetone	2.380	43	9711	8.984	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038875.D
Acq On : 17 Nov 2023 13:10
Operator : JC/MD
Sample : 05403-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB02-20231107

Quant Time: Nov 17 22:49:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

