

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038883.D
 Acq On : 17 Nov 2023 16:15
 Operator : JC/MD
 Sample : 05403-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP04-20231112

Quant Time: Nov 17 22:52:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	245553	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92122	46.760	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.520%
35) Dibromofluoromethane	5.385	113	78903	44.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.980%
50) Toluene-d8	8.647	98	312504	46.056	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.120%
62) 4-Bromofluorobenzene	11.079	95	131190	48.299	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038883.D
Acq On : 17 Nov 2023 16:15
Operator : JC/MD
Sample : 05403-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP04-20231112

Quant Time: Nov 17 22:52:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

