

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038884.D
 Acq On : 17 Nov 2023 16:39
 Operator : JC/MD
 Sample : 05404-10 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20231109

Quant Time: Nov 17 22:52:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	115636	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	235058	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88221	47.347	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.700%		
35) Dibromofluoromethane	5.385	113	74895	44.117	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.240%		
50) Toluene-d8	8.646	98	298986	46.031	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	92.060%		
62) 4-Bromofluorobenzene	11.079	95	126083	48.492	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.980%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.330	101	120657	84.519	ug/l	98
12) 1,1-Dichloroethene	2.324	96	1238	0.859	ug/l #	60
20) Methylene Chloride	2.788	84	2068	1.157	ug/l #	83
27) cis-1,2-Dichloroethene	4.495	96	900	0.486	ug/l	75
44) Trichloroethene	7.128	130	35775	20.423	ug/l	99
64) Tetrachloroethene	9.274	164	30501	20.459	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038884.D
Acq On : 17 Nov 2023 16:39
Operator : JC/MD
Sample : 05404-10 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20231109

Quant Time: Nov 17 22:52:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

