

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038889.D
 Acq On : 17 Nov 2023 18:34
 Operator : JC/MD
 Sample : 05404-33 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D2-20231110

Quant Time: Nov 17 22:54:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126783	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91780	46.523	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.040%		
35) Dibromofluoromethane	5.385	113	79031	44.403	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.800%		
50) Toluene-d8	8.647	98	310489	45.595	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	91.180%#		
62) 4-Bromofluorobenzene	11.079	95	130987	48.051	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.100%		
Target Compounds						
20) Methylene Chloride	2.788	84	2442	1.290	ug/l	87
44) Trichloroethene	7.129	130	68917	37.527	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038889.D
Acq On : 17 Nov 2023 18:34
Operator : JC/MD
Sample : 05404-33 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D2-20231110

Quant Time: Nov 17 22:54:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

