

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038890.D
 Acq On : 17 Nov 2023 18:57
 Operator : JC/MD
 Sample : PB157201TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB157201TB

Quant Time: Nov 17 22:54:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	114998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236247	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120341	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88596	47.812	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.620%
35) Dibromofluoromethane	5.385	113	74003	43.372	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.740%
50) Toluene-d8	8.647	98	297982	45.646	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.300%#
62) 4-Bromofluorobenzene	11.079	95	127837	48.919	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.840%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	25444	25.794	ug/l #	89
20) Methylene Chloride	2.788	84	6098	3.429	ug/l	95
37) Ethyl Acetate	4.733	43	3473	1.090	ug/l #	96
43) Isopropyl Acetate	6.336	43	165040	32.413	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038890.D
Acq On : 17 Nov 2023 18:57
Operator : JC/MD
Sample : PB157201TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB157201TB

Quant Time: Nov 17 22:54:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

