

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038891.D
 Acq On : 17 Nov 2023 19:21
 Operator : JC/MD
 Sample : 05466-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB21144

Quant Time: Nov 17 22:55:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	115122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237859	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129164	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89945	48.488	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.980%
35) Dibromofluoromethane	5.385	113	77042	44.847	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.700%
50) Toluene-d8	8.647	98	302455	46.017	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.040%
62) 4-Bromofluorobenzene	11.079	95	134024	50.939	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.880%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	29310	29.681	ug/l	92
20) Methylene Chloride	2.788	84	61777	34.704	ug/l	97
37) Ethyl Acetate	4.733	43	3489	1.088	ug/l #	86
43) Isopropyl Acetate	6.336	43	130690	25.493	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038891.D
Acq On : 17 Nov 2023 19:21
Operator : JC/MD
Sample : 05466-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB21144

Quant Time: Nov 17 22:55:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

