

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038899.D
 Acq On : 18 Nov 2023 01:15
 Operator : JC/MD
 Sample : 05404-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT172S1-20231109

Quant Time: Nov 18 05:04:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229746	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86852	47.629	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.260%
35) Dibromofluoromethane	5.385	113	73697	44.415	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.840%
50) Toluene-d8	8.646	98	291563	45.926	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.860%#
62) 4-Bromofluorobenzene	11.079	95	125148	49.245	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.480%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1868	1.337	ug/l	91
44) Trichloroethene	7.128	130	60980	35.617	ug/l	98
64) Tetrachloroethene	9.274	164	2757	1.882	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038899.D
Acq On : 18 Nov 2023 01:15
Operator : JC/MD
Sample : 05404-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT172S1-20231109

Quant Time: Nov 18 05:04:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

