

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038903.D
 Acq On : 18 Nov 2023 02:48
 Operator : JC/MD
 Sample : 05404-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20231109

Quant Time: Nov 18 05:06:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	114395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229260	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86679	47.024	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.040%
35) Dibromofluoromethane	5.385	113	73243	44.235	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.460%
50) Toluene-d8	8.647	98	288585	45.554	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.100%#
62) 4-Bromofluorobenzene	11.079	95	122332	48.239	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.480%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	41239	29.201	ug/l	98
12) 1,1-Dichloroethene	2.319	96	13794	9.675	ug/l	89
20) Methylene Chloride	2.794	84	1368	0.773	ug/l	97
24) 1,1-Dichloroethane	3.617	63	5645	1.928	ug/l #	90
27) cis-1,2-Dichloroethene	4.495	96	5487	2.992	ug/l	85
30) Chloroform	5.105	83	3188	1.115	ug/l	91
38) Carbon Tetrachloride	5.678	117	2458	1.121	ug/l #	95
44) Trichloroethene	7.129	130	2014454	1179.101	ug/l	94
55) 1,1,2-Trichloroethane	9.153	97	1921	1.057	ug/l	88
64) Tetrachloroethene	9.275	164	6514	4.484	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038903.D
Acq On : 18 Nov 2023 02:48
Operator : JC/MD
Sample : 05404-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20231109

Quant Time: Nov 18 05:06:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

