

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038904.D
 Acq On : 18 Nov 2023 03:11
 Operator : JC/MD
 Sample : 05404-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D2-20231109

Quant Time: Nov 18 05:06:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	106875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217984	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85095	49.413	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.820%
35) Dibromofluoromethane	5.385	113	69499	44.145	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.300%
50) Toluene-d8	8.647	98	275143	45.678	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.360%#
62) 4-Bromofluorobenzene	11.079	95	116959	48.506	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.020%
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.892	101	847	0.426	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.331	101	46604	35.322	ug/l	98
12) 1,1-Dichloroethene	2.319	96	1351	1.014	ug/l #	61
27) cis-1,2-Dichloroethene	4.501	96	4183	2.442	ug/l	87
44) Trichloroethene	7.129	130	204033	125.602	ug/l	99
64) Tetrachloroethene	9.275	164	61926	45.299	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038904.D
Acq On : 18 Nov 2023 03:11
Operator : JC/MD
Sample : 05404-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D2-20231109

Quant Time: Nov 18 05:06:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

