

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038906.D
 Acq On : 18 Nov 2023 03:58
 Operator : JC/MD
 Sample : 05404-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20231109

Quant Time: Nov 18 05:07:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227953	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86592	47.317	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.640%
35) Dibromofluoromethane	5.385	113	72895	44.277	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.560%
50) Toluene-d8	8.647	98	289444	45.951	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.900%#
62) 4-Bromofluorobenzene	11.079	95	121731	48.277	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.560%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	211725	151.005	ug/l	97
12) 1,1-Dichloroethene	2.319	96	2267	1.601	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	3461	1.901	ug/l	85
44) Trichloroethene	7.129	130	183755	108.172	ug/l	99
64) Tetrachloroethene	9.275	164	15427	10.707	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038906.D
Acq On : 18 Nov 2023 03:58
Operator : JC/MD
Sample : 05404-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20231109

Quant Time: Nov 18 05:07:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

