

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038912.D
 Acq On : 18 Nov 2023 06:17
 Operator : JC/MD
 Sample : 05404-28
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20231110

Quant Time: Nov 20 00:34:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	223566	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86568	49.004	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.000%
35) Dibromofluoromethane	5.385	113	72295	44.774	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.540%
50) Toluene-d8	8.647	98	281282	45.532	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.060%#
62) 4-Bromofluorobenzene	11.079	95	117938	47.691	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.380%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	2789	2.061	ug/l	96
20) Methylene Chloride	2.788	84	1450	0.855	ug/l	84
27) cis-1,2-Dichloroethene	4.501	96	601	0.342	ug/l	78
44) Trichloroethene	7.123	130	239770	143.917	ug/l	97
64) Tetrachloroethene	9.275	164	801	0.573	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038912.D
Acq On : 18 Nov 2023 06:17
Operator : JC/MD
Sample : 05404-28
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20231110

Quant Time: Nov 20 00:34:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

