

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038913.D
 Acq On : 18 Nov 2023 06:40
 Operator : JC/MD
 Sample : 05404-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20231109

Quant Time: Nov 20 00:34:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	103798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208501	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81522	48.741	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.480%
35) Dibromofluoromethane	5.385	113	67452	44.793	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.580%
50) Toluene-d8	8.647	98	268651	46.629	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.260%
62) 4-Bromofluorobenzene	11.079	95	114831	49.789	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.580%
Target Compounds						
					Qvalue	
30) Chloroform	5.092	83	1335	0.515	ug/l	91
44) Trichloroethene	7.129	130	8672	5.581	ug/l	86
64) Tetrachloroethene	9.268	164	894	0.662	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038913.D
Acq On : 18 Nov 2023 06:40
Operator : JC/MD
Sample : 05404-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20231109

Quant Time: Nov 20 00:34:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

