

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038868.D
 Acq On : 17 Nov 2023 09:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 17 22:45:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.583	0.605	-3.8	80	0.00
3 P	Chloromethane	0.761	0.678	10.9	77	0.00
4 C	Vinyl Chloride	0.734	0.710	3.3#	81	0.00
5 T	Bromomethane	0.303	0.381	-25.7#	91	0.00
6 T	Chloroethane	0.419	0.419	0.0	82	0.00
7 T	Trichlorofluoromethane	0.931	0.936	-0.5	84	0.00
8 T	Diethyl Ether	0.405	0.386	4.7	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.617	0.586	5.0	79	0.00
10 T	Methyl Iodide	0.751	0.694	7.6	74	0.00
11 T	Tert butyl alcohol	0.224	0.207	7.6	78	0.00
12 CM	1,1-Dichloroethene	0.623	0.584	6.3#	78	0.00
13 T	Acrolein	0.119	0.128	-7.6	97	0.00
14 T	Allyl chloride	1.248	1.149	7.9	78	0.00
15 T	Acrylonitrile	0.473	0.434	8.2	76	0.00
16 T	Acetone	0.429	0.473	-10.3	93	0.00
17 T	Carbon Disulfide	1.930	1.650	14.5	75	0.00
18 T	Methyl Acetate	2.037	1.868	8.3	76	0.00
19 T	Methyl tert-butyl Ether	2.192	2.081	5.1	79	0.00
20 T	Methylene Chloride	0.773	0.695	10.1	81	0.00
21 T	trans-1,2-Dichloroethene	0.688	0.638	7.3	79	0.00
22 T	Diisopropyl ether	2.289	2.180	4.8	79	0.00
23 T	Vinyl Acetate	1.649	1.574	4.5	77	0.00
24 P	1,1-Dichloroethane	1.280	1.197	6.5	79	0.00
25 T	2-Butanone	0.696	0.655	5.9	79	0.00
26 T	2,2-Dichloropropane	1.110	1.053	5.1	80	0.00
27 T	cis-1,2-Dichloroethene	0.802	0.745	7.1	79	0.00
28 T	Bromochloromethane	0.615	0.608	1.1	84	0.00
29 T	Tetrahydrofuran	0.442	0.412	6.8	76	0.00
30 C	Chloroform	1.250	1.190	4.8#	81	0.00
31 T	Cyclohexane	1.150	1.067	7.2	78	0.00
32 T	1,1,1-Trichloroethane	1.083	1.021	5.7	80	0.00
33 S	1,2-Dichloroethane-d4	0.806	0.764	5.2	78	-0.01
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.361	0.344	4.7	77	0.00
36 T	1,1-Dichloropropene	0.518	0.484	6.6	78	0.00
37 T	Ethyl Acetate	0.674	0.632	6.2	76	0.00
38 T	Carbon Tetrachloride	0.478	0.475	0.6	82	0.00
39 T	Methylcyclohexane	0.675	0.632	6.4	78	0.00
40 TM	Benzene	1.531	1.455	5.0	78	0.00
41 T	Methacrylonitrile	0.350	0.324	7.4	73	0.00
42 TM	1,2-Dichloroethane	0.507	0.494	2.6	80	0.00
43 T	Isopropyl Acetate	1.078	1.031	4.4	78	0.00
44 TM	Trichloroethene	0.373	0.362	2.9	79	0.00
45 C	1,2-Dichloropropane	0.412	0.399	3.2#	78	0.00
46 T	Dibromomethane	0.275	0.268	2.5	79	0.00
47 T	Bromodichloromethane	0.545	0.519	4.8	80	0.00
48 T	Methyl methacrylate	0.523	0.475	9.2	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038868.D
 Acq On : 17 Nov 2023 09:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 17 22:45:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	83	0.00
50 S	Toluene-d8	1.382	1.304	5.6	77	0.00
51 T	4-Methyl-2-Pentanone	0.689	0.657	4.6	77	0.00
52 CM	Toluene	0.950	0.912	4.0#	79	0.00
53 T	t-1,3-Dichloropropene	0.648	0.619	4.5	78	0.00
54 T	cis-1,3-Dichloropropene	0.669	0.649	3.0	78	0.00
55 T	1,1,2-Trichloroethane	0.396	0.381	3.8	79	0.00
56 T	Ethyl methacrylate	0.671	0.642	4.3	76	0.00
57 T	1,3-Dichloropropane	0.656	0.626	4.6	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.312	0.313	-0.3	96	0.00
59 T	2-Hexanone	0.552	0.537	2.7	78	0.00
60 T	Dibromochloromethane	0.403	0.391	3.0	80	0.00
61 T	1,2-Dibromoethane	0.416	0.402	3.4	79	0.00
62 S	4-Bromofluorobenzene	0.553	0.517	6.5	76	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.311	0.298	4.2	80	0.00
65 PM	Chlorobenzene	1.077	1.036	3.8	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.375	3.1	80	0.00
67 C	Ethyl Benzene	1.979	1.924	2.8#	79	0.00
68 T	m/p-Xylenes	0.749	0.723	3.5	79	0.00
69 T	o-Xylene	0.741	0.714	3.6	80	0.00
70 T	Styrene	1.251	1.200	4.1	79	0.00
71 P	Bromoform	0.323	0.319	1.2	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.896	3.662	6.0	80	0.00
74 T	N-amyl acetate	2.059	1.969	4.4	78	0.00
75 P	1,1,2,2-Tetrachloroethane	1.521	1.399	8.0	78	0.00
76 T	1,2,3-Trichloropropane	1.247	1.200	3.8	80	0.00
77 T	Bromobenzene	0.875	0.822	6.1	80	0.00
78 T	n-propylbenzene	4.795	4.576	4.6	80	0.00
79 T	2-Chlorotoluene	2.846	2.630	7.6	79	0.00
80 T	1,3,5-Trimethylbenzene	3.276	3.141	4.1	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.562	0.532	5.3	77	0.00
82 T	4-Chlorotoluene	3.281	3.093	5.7	79	0.00
83 T	tert-Butylbenzene	3.226	3.083	4.4	81	0.00
84 T	1,2,4-Trimethylbenzene	3.315	3.159	4.7	80	0.00
85 T	sec-Butylbenzene	4.220	4.034	4.4	80	0.00
86 T	p-Isopropyltoluene	3.387	3.239	4.4	81	0.00
87 T	1,3-Dichlorobenzene	1.714	1.627	5.1	80	0.00
88 T	1,4-Dichlorobenzene	1.754	1.630	7.1	80	0.00
89 T	n-Butylbenzene	3.418	3.276	4.2	80	0.00
90 T	Hexachloroethane	0.676	0.647	4.3	80	0.00
91 T	1,2-Dichlorobenzene	1.626	1.563	3.9	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.373	0.357	4.3	79	0.00
93 T	1,2,4-Trichlorobenzene	1.089	1.042	4.3	81	0.00
94 T	Hexachlorobutadiene	0.386	0.373	3.4	80	0.00
95 T	Naphthalene	4.058	3.782	6.8	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038868.D
Acq On : 17 Nov 2023 09:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 17 22:45:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.033	0.990	4.2	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6