

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038894.D
 Acq On : 17 Nov 2023 22:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 18 01:35:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	69	0.00
2 T	Dichlorodifluoromethane	0.583	0.615	-5.5	69	0.00
3 P	Chloromethane	0.761	0.705	7.4	67	0.00
4 C	Vinyl Chloride	0.734	0.741	-1.0#	71	0.00
5 T	Bromomethane	0.303	0.340	-12.2	68	0.00
6 T	Chloroethane	0.419	0.450	-7.4	74	0.00
7 T	Trichlorofluoromethane	0.931	0.981	-5.4	74	0.00
8 T	Diethyl Ether	0.405	0.416	-2.7	74	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.617	0.584	5.3	66	0.00
10 T	Methyl Iodide	0.751	0.754	-0.4	68	0.00
11 T	Tert butyl alcohol	0.224	0.253	-12.9	80	0.00
12 CM	1,1-Dichloroethene	0.623	0.590	5.3#	66	0.00
13 T	Acrolein	0.119	0.119	0.0	76	0.00
14 T	Allyl chloride	1.248	1.196	4.2	68	0.00
15 T	Acrylonitrile	0.473	0.494	-4.4	72	0.00
16 T	Acetone	0.429	0.416	3.0	69	0.00
17 T	Carbon Disulfide	1.930	1.618	16.2	62	0.00
18 T	Methyl Acetate	2.037	2.148	-5.4	73	0.00
19 T	Methyl tert-butyl Ether	2.192	2.284	-4.2	72	0.00
20 T	Methylene Chloride	0.773	0.718	7.1	70	0.00
21 T	trans-1,2-Dichloroethene	0.688	0.654	4.9	68	0.00
22 T	Diisopropyl ether	2.289	2.315	-1.1	70	0.00
23 T	Vinyl Acetate	1.649	1.667	-1.1	69	0.00
24 P	1,1-Dichloroethane	1.280	1.260	1.6	69	0.00
25 T	2-Butanone	0.696	0.716	-2.9	73	0.00
26 T	2,2-Dichloropropane	1.110	0.942	15.1	60	0.00
27 T	cis-1,2-Dichloroethene	0.802	0.791	1.4	71	0.00
28 T	Bromochloromethane	0.615	0.644	-4.7	75	0.00
29 T	Tetrahydrofuran	0.442	0.475	-7.5	73	0.00
30 C	Chloroform	1.250	1.254	-0.3#	72	0.00
31 T	Cyclohexane	1.150	1.054	8.3	65	0.00
32 T	1,1,1-Trichloroethane	1.083	1.078	0.5	71	0.00
33 S	1,2-Dichloroethane-d4	0.806	0.826	-2.5	71	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	68	0.00
35 S	Dibromofluoromethane	0.361	0.362	-0.3	69	0.00
36 T	1,1-Dichloropropene	0.518	0.492	5.0	67	0.00
37 T	Ethyl Acetate	0.674	0.703	-4.3	72	0.00
38 T	Carbon Tetrachloride	0.478	0.488	-2.1	72	0.00
39 T	Methylcyclohexane	0.675	0.607	10.1	64	0.00
40 TM	Benzene	1.531	1.492	2.5	68	0.00
41 T	Methacrylonitrile	0.350	0.369	-5.4	71	0.00
42 TM	1,2-Dichloroethane	0.507	0.533	-5.1	73	0.00
43 T	Isopropyl Acetate	1.078	1.114	-3.3	72	0.00
44 TM	Trichloroethene	0.373	0.372	0.3	70	0.00
45 C	1,2-Dichloropropane	0.412	0.412	0.0	69	0.00
46 T	Dibromomethane	0.275	0.285	-3.6	72	0.00
47 T	Bromodichloromethane	0.545	0.552	-1.3	72	0.00
48 T	Methyl methacrylate	0.523	0.519	0.8	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
 Data File : VX038894.D
 Acq On : 17 Nov 2023 22:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 18 01:35:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.012	-20.0	86	0.00
50 S	Toluene-d8	1.382	1.362	1.4	68	0.00
51 T	4-Methyl-2-Pentanone	0.689	0.739	-7.3	74	0.00
52 CM	Toluene	0.950	0.928	2.3#	69	0.00
53 T	t-1,3-Dichloropropene	0.648	0.634	2.2	68	0.00
54 T	cis-1,3-Dichloropropene	0.669	0.660	1.3	68	0.00
55 T	1,1,2-Trichloroethane	0.396	0.397	-0.3	70	0.00
56 T	Ethyl methacrylate	0.671	0.694	-3.4	70	0.00
57 T	1,3-Dichloropropane	0.656	0.670	-2.1	71	0.00
58 T	2-Chloroethyl Vinyl ether	0.312	0.334	-7.1	88	0.00
59 T	2-Hexanone	0.552	0.588	-6.5	73	0.00
60 T	Dibromochloromethane	0.403	0.421	-4.5	74	0.00
61 T	1,2-Dibromoethane	0.416	0.430	-3.4	72	0.00
62 S	4-Bromofluorobenzene	0.553	0.539	2.5	68	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	68	0.00
64 T	Tetrachloroethene	0.311	0.312	-0.3	72	0.00
65 PM	Chlorobenzene	1.077	1.056	1.9	70	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.397	-2.6	73	0.00
67 C	Ethyl Benzene	1.979	1.959	1.0#	69	0.00
68 T	m/p-Xylenes	0.749	0.739	1.3	69	0.00
69 T	o-Xylene	0.741	0.736	0.7	70	0.00
70 T	Styrene	1.251	1.235	1.3	70	0.00
71 P	Bromoform	0.323	0.338	-4.6	73	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
73 T	Isopropylbenzene	3.896	3.764	3.4	70	0.00
74 T	N-amyl acetate	2.059	2.070	-0.5	70	0.00
75 P	1,1,2,2-Tetrachloroethane	1.521	1.493	1.8	71	0.00
76 T	1,2,3-Trichloropropane	1.247	1.265	-1.4	72	0.00
77 T	Bromobenzene	0.875	0.865	1.1	72	0.00
78 T	n-propylbenzene	4.795	4.638	3.3	69	0.00
79 T	2-Chlorotoluene	2.846	2.706	4.9	70	0.00
80 T	1,3,5-Trimethylbenzene	3.276	3.235	1.3	71	0.00
81 T	trans-1,4-Dichloro-2-butene	0.562	0.528	6.0	66	0.00
82 T	4-Chlorotoluene	3.281	3.172	3.3	70	0.00
83 T	tert-Butylbenzene	3.226	3.119	3.3	70	0.00
84 T	1,2,4-Trimethylbenzene	3.315	3.267	1.4	71	0.00
85 T	sec-Butylbenzene	4.220	4.163	1.4	71	0.00
86 T	p-Isopropyltoluene	3.387	3.310	2.3	71	0.00
87 T	1,3-Dichlorobenzene	1.714	1.669	2.6	71	0.00
88 T	1,4-Dichlorobenzene	1.754	1.676	4.4	70	0.00
89 T	n-Butylbenzene	3.418	3.275	4.2	69	0.00
90 T	Hexachloroethane	0.676	0.659	2.5	70	0.00
91 T	1,2-Dichlorobenzene	1.626	1.634	-0.5	73	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.373	0.383	-2.7	73	0.00
93 T	1,2,4-Trichlorobenzene	1.089	1.073	1.5	71	0.00
94 T	Hexachlorobutadiene	0.386	0.374	3.1	69	0.00
95 T	Naphthalene	4.058	4.065	-0.2	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111723\
Data File : VX038894.D
Acq On : 17 Nov 2023 22:30
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 18 01:35:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.033	1.044	-1.1	72	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5