

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111725\
 Data File : VX048618.D
 Acq On : 17 Nov 2025 13:32
 Operator : JC/MD
 Sample : Q3638-01
 Misc : 3.44g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MOO-25-0327

Quant Time: Nov 17 13:50:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	131451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	328491	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	296346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	165000	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	119556	65.272	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	130.540%#
35) Dibromofluoromethane	5.367	113	95512	38.416	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	76.840%
50) Toluene-d8	8.635	98	345155	44.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.020%#
62) 4-Bromofluorobenzene	11.067	95	173870	60.169	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	120.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111725\
Data File : VX048618.D
Acq On : 17 Nov 2025 13:32
Operator : JC/MD
Sample : Q3638-01
Misc : 3.44g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MOO-25-0327

Quant Time: Nov 17 13:50:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

