

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032916.D
 Acq On : 18 Nov 2022 11:40
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050

Quant Time: Nov 19 03:02:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	90489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	140213	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69642	55.389	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.780%
35) Dibromofluoromethane	5.385	113	58801	58.792	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	117.580%
50) Toluene-d8	8.647	98	186229	53.549	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.100%
62) 4-Bromofluorobenzene	11.079	95	79484	58.509	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.020%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	44893	43.015	ug/l	95
3) Chloromethane	1.294	50	37217	42.167	ug/l	98
4) Vinyl Chloride	1.374	62	43344	44.622	ug/l	96
5) Bromomethane	1.618	94	47138	60.724	ug/l	96
6) Chloroethane	1.691	64	39471	60.570	ug/l	99
7) Trichlorofluoromethane	1.892	101	91863	49.911	ug/l	99
8) Diethyl Ether	2.130	74	31093	52.130	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.331	101	52562	54.107	ug/l	100
10) Methyl Iodide	2.453	142	56511	52.212	ug/l	99
11) Tert butyl alcohol	2.983	59	38074	208.052	ug/l #	85
12) 1,1-Dichloroethene	2.319	96	44667	49.221	ug/l	97
13) Acrolein	2.233	56	61090	334.389	ug/l	99
14) Allyl chloride	2.666	41	76207	52.680	ug/l	100
15) Acrylonitrile	3.062	53	124174	250.021	ug/l	100
16) Acetone	2.386	43	136863	299.496	ug/l	98
17) Carbon Disulfide	2.514	76	89280	41.305	ug/l	97
18) Methyl Acetate	2.703	43	83567	56.235	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	185155	58.051	ug/l	100
20) Methylene Chloride	2.788	84	54812	52.313	ug/l	96
21) trans-1,2-Dichloroethene	3.093	96	49014	49.620	ug/l	99
22) Diisopropyl ether	3.757	45	185656	59.334	ug/l	97
23) Vinyl Acetate	3.715	43	690728	273.235	ug/l	100
24) 1,1-Dichloroethane	3.605	63	99820	55.462	ug/l	99
25) 2-Butanone	4.556	43	182966	258.431	ug/l	97
26) 2,2-Dichloropropane	4.471	77	84731	59.224	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	64083	56.125	ug/l	98
28) Bromochloromethane	4.891	49	31730	49.495	ug/l	99
29) Tetrahydrofuran	5.001	42	106764	236.487	ug/l	98
30) Chloroform	5.086	83	116292	58.885	ug/l	98
31) Cyclohexane	5.471	56	74924	49.147	ug/l	97
32) 1,1,1-Trichloroethane	5.385	97	102931	58.081	ug/l	100
36) 1,1-Dichloropropene	5.690	75	74147	53.237	ug/l	99
37) Ethyl Acetate	4.708	43	70518	48.642	ug/l	99
38) Carbon Tetrachloride	5.678	117	88239	57.121	ug/l	99
39) Methylcyclohexane	7.379	83	83447	50.640	ug/l	98
40) Benzene	6.037	78	215310	54.630	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032916.D
 Acq On : 18 Nov 2022 11:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Nov 19 03:02:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	40559	54.919	ug/l	98
42) 1,2-Dichloroethane	6.080	62	94531	58.675	ug/l	100
43) Isopropyl Acetate	6.336	43	121728	54.686	ug/l	99
44) Trichloroethene	7.123	130	61152	55.117	ug/l	98
45) 1,2-Dichloropropane	7.427	63	58778	58.097	ug/l	98
46) Dibromomethane	7.580	93	44949	57.931	ug/l	99
47) Bromodichloromethane	7.818	83	93742	63.746	ug/l	97
48) Methyl methacrylate	7.690	41	62082	54.002	ug/l	99
49) 1,4-Dioxane	7.690	88	19916	909.045	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	374988	267.454	ug/l	99
52) Toluene	8.720	92	145263	56.492	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	93345	62.213	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	97060	60.481	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	65857	60.814	ug/l	98
56) Ethyl methacrylate	9.116	69	92852	59.581	ug/l	99
57) 1,3-Dichloropropane	9.305	76	103299	59.068	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	219030	291.888	ug/l	99
59) 2-Hexanone	9.427	43	280844	269.952	ug/l	99
60) Dibromochloromethane	9.519	129	74927	63.345	ug/l	99
61) 1,2-Dibromoethane	9.610	107	67814	59.100	ug/l	100
64) Tetrachloroethene	9.275	164	52679	56.412	ug/l	97
65) Chlorobenzene	10.079	112	163862	57.336	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	66548	61.870	ug/l	100
67) Ethyl Benzene	10.195	91	292807	57.293	ug/l	99
68) m/p-Xylenes	10.299	106	228713	115.249	ug/l	99
69) o-Xylene	10.640	106	113602	58.877	ug/l	99
70) Styrene	10.653	104	191018	59.569	ug/l	100
71) Bromoform	10.799	173	53410	62.803	ug/l #	97
73) Isopropylbenzene	10.963	105	309491	54.773	ug/l	100
74) N-amyl acetate	10.842	43	113654	53.235	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.213	83	98424	54.267	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	102134	64.382	ug/l	88
77) Bromobenzene	11.195	156	76583	55.284	ug/l	99
78) n-propylbenzene	11.305	91	356321	54.734	ug/l	100
79) 2-Chlorotoluene	11.366	91	214839	55.105	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	267340	55.359	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	26970	54.097	ug/l	99
82) 4-Chlorotoluene	11.457	91	252886	55.433	ug/l	99
83) tert-Butylbenzene	11.713	119	271267	56.264	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	265995	55.380	ug/l	100
85) sec-Butylbenzene	11.890	105	321629	55.441	ug/l	99
86) p-Isopropyltoluene	12.012	119	276744	56.219	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	146707	56.144	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	148327	54.716	ug/l	99
89) n-Butylbenzene	12.335	91	239177	55.849	ug/l	100
90) Hexachloroethane	12.536	117	45942	57.138	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	148447	57.615	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	19672	49.469	ug/l	100
93) 1,2,4-Trichlorobenzene	13.591	180	92657	55.153	ug/l	98
94) Hexachlorobutadiene	13.725	225	39709	56.595	ug/l	98
95) Naphthalene	13.774	128	291336	52.044	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	93175	54.560	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032916.D
Acq On : 18 Nov 2022 11:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: Nov 19 03:02:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 90489
Ion Ratio Lower Upper
168 100
99 56.4 46.1 69.1



#2
Dichlorodifluoromethane
Concen: 43.015 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 85 Resp: 44893
Ion Ratio Lower Upper
85 100
87 34.0 15.6 46.7









#3
Chloromethane
Concen: 42.167 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 37217
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2



#4
Vinyl Chloride
Concen: 44.622 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 62 Resp: 43344
Ion Ratio Lower Upper
62 100
64 30.1 25.8 38.6





#5

Bromomethane

Concen: 60.724 ug/l

RT: 1.618 min Scan# 81

Delta R.T. 0.012 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 94 Resp: 47138

Ion Ratio Lower Upper

94 100

96 93.6 77.6 116.4



#6

Chloroethane

Concen: 60.570 ug/l

RT: 1.691 min Scan# 99

Delta R.T. 0.006 min

Lab File: VX032916.D




Acq: 18 Nov 2022 11:40


Tgt Ion: 64 Resp: 39471

Ion Ratio Lower Upper

64 100

66 33.0 26.0 39.0









#7
 Trichlorofluoromethane
 Concen: 49.911 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 91863
 Ion Ratio Lower Upper
 101 100
 103 63.5 51.4 77.2



#8
 Diethyl Ether
 Concen: 52.130 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40

Tgt Ion: 74 Resp: 31093
 Ion Ratio Lower Upper
 74 100
 45 92.1 46.6 139.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 54.107 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 52562
Ion Ratio Lower Upper
101 100
85 46.1 37.0 55.4
151 81.0 64.4 96.6



#10
Methyl Iodide
Concen: 52.212 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion:142 Resp: 56511
Ion Ratio Lower Upper
142 100
127 49.2 38.6 57.8
141 14.1 11.4 17.2









#11
Tert butyl alcohol
Concen: 208.052 ug/l
RT: 2.983 min Scan# 311
Delta R.T. 0.018 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 38074
Ion Ratio Lower Upper
59 100
57 4.8 8.2 12.4#



#12
1,1-Dichloroethene
Concen: 49.221 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 96 Resp: 44667
Ion Ratio Lower Upper
96 100
61 161.7 125.1 187.7
98 63.8 50.6 76.0





#13

Acrolein

Concen: 334.389 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 56 Resp: 61090

Ion Ratio Lower Upper

56 100

55 69.9 56.9 85.3



#14

Allyl chloride

Concen: 52.680 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40




Tgt Ion: 41 Resp: 76207


Ion Ratio Lower Upper

41 100

39 76.4 61.4 92.2

76 36.2 29.2 43.8









#15
Acrylonitrile
Concen: 250.021 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 124174
Ion Ratio Lower Upper
53 100
52 84.3 67.4 101.0
51 37.3 30.3 45.5



#16
Acetone
Concen: 299.496 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 43 Resp: 136863
Ion Ratio Lower Upper
43 100
58 28.7 23.8 35.8









#17
Carbon Disulfide
Concen: 41.305 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 89280
Ion Ratio Lower Upper
76 100
78 8.4 7.5 11.3



#18
Methyl Acetate
Concen: 56.235 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 43 Resp: 83567
Ion Ratio Lower Upper
43 100
74 24.6 19.4 29.0









#19
Methyl tert-butyl Ether
Concen: 58.051 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 185155
Ion Ratio Lower Upper
73 100
57 20.7 16.6 25.0



#20
Methylene Chloride
Concen: 52.313 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 84 Resp: 54812
Ion Ratio Lower Upper
84 100
49 121.3 101.0 151.6
51 36.3 31.1 46.7
86 63.3 52.3 78.5









#21
trans-1,2-Dichloroethene
Concen: 49.620 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 49014
Ion Ratio Lower Upper
96 100
61 135.1 106.6 160.0
98 64.5 51.0 76.6



#22
Diisopropyl ether
Concen: 59.334 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 45 Resp: 185656
Ion Ratio Lower Upper
45 100
43 74.7 56.4 84.6
87 30.7 24.4 36.6
59 12.5 9.5 14.3









#23
 Vinyl Acetate
 Concen: 273.235 ug/l
 RT: 3.715 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 690728
 Ion Ratio Lower Upper
 43 100
 86 10.9 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 55.462 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40

Tgt Ion: 63 Resp: 99820
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.5 10.4
 100 4.3 2.3 6.8









#25
2-Butanone
Concen: 258.431 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 182966
Ion Ratio Lower Upper
43 100
72 26.2 22.2 33.2



#26
2,2-Dichloropropane
Concen: 59.224 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 77 Resp: 84731
Ion Ratio Lower Upper
77 100
97 22.4 11.4 34.1









#27
 cis-1,2-Dichloroethene
 Concen: 56.125 ug/l
 RT: 4.483 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 96 Resp: 64083
 Ion Ratio Lower Upper
 96 100
 61 145.5 0.0 286.0
 98 66.4 0.0 129.6



#28
 Bromochloromethane
 Concen: 49.495 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.012 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40

Tgt Ion: 49 Resp: 31730
 Ion Ratio Lower Upper
 49 100
 129 2.3 0.0 3.8
 130 85.4 67.5 101.3









#29
Tetrahydrofuran
Concen: 236.487 ug/l
RT: 5.001 min Scan# 642
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 42 Resp: 106764
Ion Ratio Lower Upper
42 100
72 48.7 37.6 56.4
71 44.6 34.7 52.1



#30
Chloroform
Concen: 58.885 ug/l
RT: 5.086 min Scan# 656
Delta R.T. -0.012 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 83 Resp: 116292
Ion Ratio Lower Upper
83 100
85 65.2 51.1 76.7









#31
Cyclohexane
Concen: 49.147 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 74924
Ion Ratio Lower Upper
56 100
69 31.6 24.1 36.1
84 86.8 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 58.081 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 97 Resp: 102931
Ion Ratio Lower Upper
97 100
99 64.0 51.4 77.0
61 43.5 35.2 52.8





#33

1,2-Dichloroethane-d4

Concen: 55.389 ug/l

RT: 5.952 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 65 Resp: 69642

Ion Ratio Lower Upper

65 100

67 49.9 0.0 101.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.006 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40




Tgt Ion: 114 Resp: 140213


Ion Ratio Lower Upper

114 100

63 19.9 0.0 41.6

88 16.4 0.0 33.4









#35
Dibromofluoromethane
Concen: 58.792 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 58801
Ion Ratio Lower Upper
113 100
111 102.0 82.4 123.6
192 18.9 15.4 23.2



#36
1,1-Dichloropropene
Concen: 53.237 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 75 Resp: 74147
Ion Ratio Lower Upper
75 100
110 36.0 18.4 55.2
77 30.5 24.5 36.7





#37

Ethyl Acetate

Concen: 48.642 ug/l

RT: 4.708 min Scan# 594

Delta R.T. -0.006 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 70518

Ion Ratio Lower Upper

43 100

61 13.8 11.5 17.3

70 11.0 9.1 13.7



#38

Carbon Tetrachloride

Concen: 57.121 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40




Tgt Ion: 117 Resp: 88239


Ion Ratio Lower Upper

117 100

119 95.8 77.2 115.8

121 30.3 25.8 38.6









#39
Methylcyclohexane
Concen: 50.640 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 83447
Ion Ratio Lower Upper
83 100
55 74.4 58.4 87.6
98 47.4 37.2 55.8



#40
Benzene
Concen: 54.630 ug/l
RT: 6.037 min Scan# 812
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 78 Resp: 215310
Ion Ratio Lower Upper
78 100
77 23.4 19.0 28.4









#41
Methacrylonitrile
Concen: 54.919 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	40559
Ion	Ratio	Lower	Upper
41	100		
39	60.6	47.5	71.3
67	76.9	60.5	90.7
52	32.8	27.0	40.6



#42
1,2-Dichloroethane
Concen: 58.675 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion:	62	Resp:	94531
Ion	Ratio	Lower	Upper
62	100		
98	8.2	0.0	16.0









#43
Isopropyl Acetate
Concen: 54.686 ug/l
RT: 6.336 min Scan# 80
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 121728
Ion Ratio Lower Upper
43 100
61 19.3 15.4 23.0
87 13.9 10.8 16.2



#44
Trichloroethene
Concen: 55.117 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion:130 Resp: 61152
Ion Ratio Lower Upper
130 100
95 99.4 0.0 195.2









#45
1,2-Dichloropropane
Concen: 58.097 ug/l
RT: 7.427 min Scan# 1040
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 63 Resp: 58778
Ion Ratio Lower Upper
63 100
65 29.6 24.6 36.8



#46
Dibromomethane
Concen: 57.931 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 93 Resp: 44949
Ion Ratio Lower Upper
93 100
95 82.8 65.8 98.8
174 98.3 79.8 119.6









#47
Bromodichloromethane
Concen: 63.746 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 93742
Ion Ratio Lower Upper
83 100
85 62.4 52.2 78.4
127 8.5 7.4 11.2



#48
Methyl methacrylate
Concen: 54.002 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 41 Resp: 62082
Ion Ratio Lower Upper
41 100
69 82.2 65.1 97.7
39 64.6 51.9 77.9





#49

1,4-Dioxane

Concen: 909.045 ug/l

RT: 7.690 min Scan# 1083

Delta R.T. 0.018 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 19916

Ion Ratio Lower Upper

88 100

43 34.6 26.6 39.8

58 65.5 51.7 77.5



#50

Toluene-d8

Concen: 53.549 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX032916.D




Acq: 18 Nov 2022 11:40


Tgt Ion: 98 Resp: 186229

Ion Ratio Lower Upper

98 100

100 64.1 50.9 76.3









#51
4-Methyl-2-Pentanone
Concen: 267.454 ug/l
RT: 8.574 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 374988
Ion Ratio Lower Upper
43 100
58 39.3 30.9 46.3



#52
Toluene
Concen: 56.492 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 92 Resp: 145263
Ion Ratio Lower Upper
92 100
91 174.5 138.3 207.5









#53
 t-1,3-Dichloropropene
 Concen: 62.213 ug/l
 RT: 8.976 min Scan# 1129
 Delta R.T. -0.006 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 75 Resp: 93345
 Ion Ratio Lower Upper
 75 100
 77 31.7 25.1 37.7



#54
 cis-1,3-Dichloropropene
 Concen: 60.481 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40

Tgt Ion: 75 Resp: 97060
 Ion Ratio Lower Upper
 75 100
 77 31.5 25.1 37.7
 39 52.5 45.5 68.3









#55
 1,1,2-Trichloroethane
 Concen: 60.814 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:	97	Resp:	65857
Ion	Ratio	Lower	Upper
97	100		
83	84.8	68.1	102.1
85	55.7	43.3	64.9
99	65.8	50.4	75.6



#56
 Ethyl methacrylate
 Concen: 59.581 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40

Tgt Ion:	69	Resp:	92852
Ion	Ratio	Lower	Upper
69	100		
41	67.6	54.3	81.5
39	42.4	35.2	52.8





#57

1,3-Dichloropropane

Concen: 59.068 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 103299

Ion Ratio Lower Upper

76 100

78 32.9 26.0 39.0



#58

2-Chloroethyl Vinyl ether

Concen: 291.888 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX032916.D




Acq: 18 Nov 2022 11:40


Tgt Ion: 63 Resp: 219030

Ion Ratio Lower Upper

63 100

106 29.5 23.3 34.9









#59
2-Hexanone
Concen: 269.952 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 280844
Ion Ratio Lower Upper
43 100
58 53.4 26.5 79.5



#60
Dibromochloromethane
Concen: 63.345 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 129 Resp: 74927
Ion Ratio Lower Upper
129 100
127 78.5 38.6 115.8









#61
1,2-Dibromoethane
Concen: 59.100 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 67814
Ion Ratio Lower Upper
107 100
109 94.7 76.0 114.0



#62
4-Bromofluorobenzene
Concen: 58.509 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 95 Resp: 79484
Ion Ratio Lower Upper
95 100
174 81.2 0.0 164.0
176 79.8 0.0 154.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 127584
Ion Ratio Lower Upper
117 100
82 54.2 44.3 66.5
119 31.5 25.7 38.5



#64
Tetrachloroethene
Concen: 56.412 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion:164 Resp: 52679
Ion Ratio Lower Upper
164 100
166 128.0 105.4 158.2
129 94.9 78.2 117.2
131 93.3 72.5 108.7









#65
Chlorobenzene
Concen: 57.336 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 163862
Ion Ratio Lower Upper
112 100
114 31.9 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 61.870 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion:131 Resp: 66548
Ion Ratio Lower Upper
131 100
133 95.8 47.6 142.9
119 64.5 32.4 97.0









#67
Ethyl Benzene
Concen: 57.293 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 292807
Ion Ratio Lower Upper
91 100
106 31.0 25.0 37.6



#68
m/p-Xylenes
Concen: 115.249 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 106 Resp: 228713
Ion Ratio Lower Upper
106 100
91 204.6 164.5 246.7









#69
o-Xylene
Concen: 58.877 ug/l
RT: 10.640 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 113602
Ion Ratio Lower Upper
106 100
91 217.6 109.2 327.6



#70
Styrene
Concen: 59.569 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion:104 Resp: 191018
Ion Ratio Lower Upper
104 100
78 52.3 41.8 62.8
103 56.1 44.5 66.7





#71

Bromoform

Concen: 62.803 ug/l

RT: 10.799 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:173 Resp: 53410

Ion Ratio Lower Upper

173 100

175 49.6 23.9 71.8

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40




Tgt Ion:152 Resp: 71306


Ion Ratio Lower Upper

152 100

115 91.6 44.0 132.0

150 180.3 0.0 351.8









#73
Isopropylbenzene
Concen: 54.773 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 309491
Ion Ratio Lower Upper
105 100
120 26.2 13.1 39.3



#74
N-ethyl acetate
Concen: 53.235 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 43 Resp: 113654
Ion Ratio Lower Upper
43 100
70 41.7 33.6 50.4
55 24.4 19.4 29.2
61 23.7 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 54.267 ug/l

RT: 11.213 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 98424

Ion Ratio Lower Upper

83 100

131 10.6 5.4 16.2

85 64.1 31.9 95.9



#76

1,2,3-Trichloropropane

Concen: 64.382 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.006 min

Lab File: VX032916.D




Acq: 18 Nov 2022 11:40


Tgt Ion: 75 Resp: 102134

Ion Ratio Lower Upper

75 100

77 32.4 20.0 60.0









#77
Bromobenzene
Concen: 55.284 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.006 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 76583
Ion Ratio Lower Upper
156 100
77 153.3 76.1 228.4
158 96.5 48.8 146.3



#78
n-propylbenzene
Concen: 54.734 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 91 Resp: 356321
Ion Ratio Lower Upper
91 100
120 23.1 11.6 34.8









#79
2-Chlorotoluene
Concen: 55.105 ug/l
RT: 11.366 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 214839
Ion Ratio Lower Upper
91 100
126 34.0 17.1 51.2



#80
1,3,5-Trimethylbenzene
Concen: 55.359 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 105 Resp: 267340
Ion Ratio Lower Upper
105 100
120 47.7 23.6 71.0









#81
trans-1,4-Dichloro-2-butene
Concen: 54.097 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 26970
Ion Ratio Lower Upper
75 100
53 107.5 86.9 130.3
89 49.3 38.3 57.5



#82
4-Chlorotoluene
Concen: 55.433 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion: 91 Resp: 252886
Ion Ratio Lower Upper
91 100
126 30.2 14.8 44.4









#83
 tert-Butylbenzene
 Concen: 56.264 ug/l
 RT: 11.713 min Scan# 1743
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:119 Resp: 271267
 Ion Ratio Lower Upper
 119 100
 91 57.5 29.4 88.2
 134 21.5 11.0 33.0



#84
 1,2,4-Trimethylbenzene
 Concen: 55.380 ug/l
 RT: 11.750 min Scan# 1749
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40

Tgt Ion:105 Resp: 265995
 Ion Ratio Lower Upper
 105 100
 120 43.4 21.9 65.5









#85
 sec-Butylbenzene
 Concen: 55.441 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 321629
 Ion Ratio Lower Upper
 105 100
 134 20.2 9.9 29.7



#86
 p-Isopropyltoluene
 Concen: 56.219 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40

Tgt Ion:119 Resp: 276744
 Ion Ratio Lower Upper
 119 100
 134 26.1 12.9 38.7
 91 25.0 12.5 37.5





#87

1,3-Dichlorobenzene

Concen: 56.144 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 146707

Ion Ratio Lower Upper

146 100

111 40.6 20.6 61.8

148 63.9 31.7 95.1



#88

1,4-Dichlorobenzene

Concen: 54.716 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40




Tgt Ion:146 Resp: 148327


Ion Ratio Lower Upper

146 100

111 39.9 20.5 61.6

148 63.7 32.2 96.6





#89

n-Butylbenzene

Concen: 55.849 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 91 Resp: 239177

Ion Ratio Lower Upper

91 100

92 52.5 26.2 78.6

134 26.0 12.9 38.6



#90

Hexachloroethane

Concen: 57.138 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.006 min

Lab File: VX032916.D




Acq: 18 Nov 2022 11:40


Tgt Ion: 117 Resp: 45942

Ion Ratio Lower Upper

117 100

201 74.9 37.6 112.9









#91
 1,2-Dichlorobenzene
 Concen: 57.615 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:146 Resp: 148447
 Ion Ratio Lower Upper
 146 100
 111 41.4 21.5 64.5
 148 63.4 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 49.469 ug/l
 RT: 12.945 min Scan# 1945
 Delta R.T. 0.000 min
 Lab File: VX032916.D
 Acq: 18 Nov 2022 11:40

Tgt Ion: 75 Resp: 19672
 Ion Ratio Lower Upper
 75 100
 155 83.6 41.6 124.8
 157 107.9 54.1 162.4









#93
1,2,4-Trichlorobenzene
Concen: 55.153 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 92657
Ion Ratio Lower Upper
180 100
182 94.6 48.4 145.0
145 31.4 16.2 48.5



#94
Hexachlorobutadiene
Concen: 56.595 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX032916.D
Acq: 18 Nov 2022 11:40

Tgt Ion:225 Resp: 39709
Ion Ratio Lower Upper
225 100
223 62.0 31.9 95.9
227 64.3 31.9 95.5





#95

Naphthalene

Concen: 52.044 ug/l

RT: 13.774 min Scan# 2082

Delta R.T. -0.006 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:128 Resp: 291336

Ion Ratio Lower Upper

128 100

127 12.7 10.2 15.4

129 10.8 8.9 13.3



#96

1,2,3-Trichlorobenzene

Concen: 54.560 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.000 min

Lab File: VX032916.D

Acq: 18 Nov 2022 11:40

Tgt Ion:180 Resp: 93175

Ion Ratio Lower Upper

180 100

182 94.4 47.5 142.6

145 33.3 16.9 50.6

