

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032920.D
 Acq On : 18 Nov 2022 13:23
 Operator : JC/MD
 Sample : N5671-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D2-20221115DL

Quant Time: Nov 19 03:04:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	192399	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82207	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82767	51.675	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.340%
35) Dibromofluoromethane	5.385	113	66243	48.268	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.540%
50) Toluene-d8	8.647	98	238498	49.977	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.960%
62) 4-Bromofluorobenzene	11.079	95	93991	50.421	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.840%
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	5195	3.572	ug/l	91
44) Trichloroethene	7.122	130	49003	32.187	ug/l	99
64) Tetrachloroethene	9.274	164	4058	3.183	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032920.D
Acq On : 18 Nov 2022 13:23
Operator : JC/MD
Sample : N5671-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D2-20221115DL

Quant Time: Nov 19 03:04:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

