

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032922.D
 Acq On : 18 Nov 2022 14:09
 Operator : JC/MD
 Sample : N5671-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW17811-20221115DL

Quant Time: Nov 19 03:05:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	106857	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177866	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76858	51.765	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.520%
35) Dibromofluoromethane	5.385	113	60903	48.003	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.000%
50) Toluene-d8	8.647	98	222468	50.427	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.860%
62) 4-Bromofluorobenzene	11.079	95	83770	48.610	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.220%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	400	0.323	ug/l #	74
27) cis-1,2-Dichloroethene	4.495	96	4982	3.695	ug/l	87
44) Trichloroethene	7.129	130	59086	41.981	ug/l	98
45) 1,2-Dichloropropane	7.427	63	3268	2.546	ug/l	95
64) Tetrachloroethene	9.275	164	3639	3.083	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032922.D
Acq On : 18 Nov 2022 14:09
Operator : JC/MD
Sample : N5671-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW17811-20221115DL

Quant Time: Nov 19 03:05:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

