

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032923.D
 Acq On : 18 Nov 2022 14:32
 Operator : JC/MD
 Sample : N5671-08DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20221115DL

Quant Time: Nov 19 03:05:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	172520	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67527	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74970	51.288	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.580%	
35) Dibromofluoromethane	5.391	113	59619	48.447	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.900%	
50) Toluene-d8	8.647	98	216937	50.697	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.400%	
62) 4-Bromofluorobenzene	11.079	95	80469	48.142	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.280%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	67538	59.801	ug/l	98
44) Trichloroethene	7.129	130	22233	16.286	ug/l	99
64) Tetrachloroethene	9.275	164	12281	10.944	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032923.D
Acq On : 18 Nov 2022 14:32
Operator : JC/MD
Sample : N5671-08DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20221115DL

Quant Time: Nov 19 03:05:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

