

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032926.D
 Acq On : 18 Nov 2022 15:42
 Operator : JC/MD
 Sample : N5672-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20221116DL

Quant Time: Nov 19 03:07:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	103510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	174247	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75906	52.777	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.560%
35) Dibromofluoromethane	5.385	113	60680	48.820	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.640%
50) Toluene-d8	8.647	98	217095	50.231	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.460%
62) 4-Bromofluorobenzene	11.079	95	83442	49.425	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.860%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1061	0.955	ug/l	95
44) Trichloroethene	7.123	130	84106	60.999	ug/l	97
64) Tetrachloroethene	9.281	164	666	0.572	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032926.D
Acq On : 18 Nov 2022 15:42
Operator : JC/MD
Sample : N5672-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20221116DL

Quant Time: Nov 19 03:07:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

