

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032928.D
 Acq On : 18 Nov 2022 16:28
 Operator : JC/MD
 Sample : N5671-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178I-20221115

Quant Time: Nov 19 03:07:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	104110	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173551	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75745	52.361	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.720%
35) Dibromofluoromethane	5.385	113	60113	48.558	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.120%
50) Toluene-d8	8.647	98	218726	50.812	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.079	95	84053	49.987	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.980%
Target Compounds						Qvalue
44) Trichloroethene	7.129	130	1517	1.105	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032928.D
Acq On : 18 Nov 2022 16:28
Operator : JC/MD
Sample : N5671-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178I-20221115

Quant Time: Nov 19 03:07:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

