

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032932.D
 Acq On : 18 Nov 2022 18:00
 Operator : JC/MD
 Sample : N5672-02 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D2-20221116

Quant Time: Nov 19 03:09:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103294	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	172310	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	157186	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73690	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74833	52.140	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.280%
35) Dibromofluoromethane	5.391	113	59672	48.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.100%
50) Toluene-d8	8.647	98	217152	50.809	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.079	95	83549	50.045	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.100%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.123	130	129322	94.847	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032932.D
Acq On : 18 Nov 2022 18:00
Operator : JC/MD
Sample : N5672-02 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D2-20221116

Quant Time: Nov 19 03:09:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

