

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032935.D
 Acq On : 18 Nov 2022 19:10
 Operator : JC/MD
 Sample : N5672-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20221116

Quant Time: Nov 19 03:11:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/21/2022
 Supervised By : Mahesh Dadoda 11/21/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77249	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81180	51.516	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.040%			
35) Dibromofluoromethane	5.385	113	65095	48.326	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.660%			
50) Toluene-d8	8.647	98	237407	50.686	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.380%			
62) 4-Bromofluorobenzene	11.079	95	88797	48.533	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.060%			
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	27687	22.741	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	3691	2.579	ug/l	92
30) Chloroform	5.105	83	1146m	0.463	ug/l	
44) Trichloroethene	7.129	130	801552	536.415	ug/l	99
52) Toluene	8.726	92	1802	0.520	ug/l	90
64) Tetrachloroethene	9.275	164	6808	5.455	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032935.D
Acq On : 18 Nov 2022 19:10
Operator : JC/MD
Sample : N5672-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20221116

Quant Time: Nov 19 03:11:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/21/2022
Supervised By :Mahesh Dadoda 11/21/2022

