

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
 Data File : VX032918.D
 Acq On : 18 Nov 2022 12:33
 Operator : JC/MD
 Sample : VX1118WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1118WBL01

Quant Time: Nov 19 03:03:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178004	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159119	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75498	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76476	52.065	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.120%
35) Dibromofluoromethane	5.385	113	61483	48.422	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	8.647	98	221600	50.191	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.380%
62) 4-Bromofluorobenzene	11.079	95	84501	48.996	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.000%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111822\
Data File : VX032918.D
Acq On : 18 Nov 2022 12:33
Operator : JC/MD
Sample : VX1118WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1118WBL01

Quant Time: Nov 19 03:03:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 711
Delta R.T. -0.000 min
Lab File: VX032918.D
Acq: 18 Nov 2022 12:33




Instrument :
MSVOA_X
ClientSampleId :
VX1118WBL01


Tgt Ion:168 Resp: 105713
Ion Ratio Lower Upper
168 100
99 57.1 46.1 69.1



#33
1,2-Dichloroethane-d4
Concen: 52.065 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX032918.D
Acq: 18 Nov 2022 12:33

Tgt Ion: 65 Resp: 76476
Ion Ratio Lower Upper
65 100
67 50.7 0.0 101.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 91

Delta R.T. 0.000 min

Lab File: VX032918.D

Acq: 18 Nov 2022 12:33

Instrument :

MSVOA_X

ClientSampleId :

VX1118WBL01





Tgt Ion:114 Resp: 178004

Ion Ratio Lower Upper

114 100

63 19.6 0.0 41.6

88 15.4 0.0 33.4



#35

Dibromofluoromethane

Concen: 48.422 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.000 min

Lab File: VX032918.D

Acq: 18 Nov 2022 12:33




Tgt Ion:113 Resp: 61483


Ion Ratio Lower Upper

113 100

111 101.8 82.4 123.6

192 19.3 15.4 23.2









#50
Toluene-d8
Concen: 50.191 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX032918.D
Acq: 18 Nov 2022 12:33




Instrument :
MSVOA_X
ClientSampleId :
VX1118WBL01


Tgt Ion: 98 Resp: 221600
Ion Ratio Lower Upper
98 100
100 63.9 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 48.996 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX032918.D
Acq: 18 Nov 2022 12:33

Tgt Ion: 95 Resp: 84501
Ion Ratio Lower Upper
95 100
174 83.0 0.0 164.0
176 80.1 0.0 154.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX032918.D
Acq: 18 Nov 2022 12:33

Instrument :
MSVOA_X
ClientSampleId :
VX1118WBL01

Tgt Ion:117 Resp: 159119
Ion Ratio Lower Upper
117 100
82 55.6 44.3 66.5
119 32.4 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX032918.D
Acq: 18 Nov 2022 12:33

Tgt Ion:152 Resp: 75498
Ion Ratio Lower Upper
152 100
115 60.7 44.0 132.0
150 155.5 0.0 351.8

