

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111918\
 Data File : VX006089.D
 Acq On : 19 Nov 2018 14:21
 Operator : JC/MD
 Sample : J5767-11
 Misc : 6.61g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-111418-B

Quant Time: Nov 19 15:45:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	135179	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	199597	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	156494	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	86621	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	73004	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
35) Dibromofluoromethane	5.51	113	62099	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
50) Toluene-d8	8.71	98	202594	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	
62) 4-Bromofluorobenzene	11.14	95	75631	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
Target Compounds						
52) Toluene	8.78	92	3314	1.06	ug/l	98
68) m/p-Xylenes	10.36	106	3033	1.58	ug/l	99
69) o-Xylene	10.70	106	2098	1.22	ug/l	94
84) 1,2,4-Trimethylbenzene	11.81	105	19001	3.93	ug/l	100
95) Naphthalene	13.83	128	618483	98.09	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111918\
Data File : VX006089.D
Acq On : 19 Nov 2018 14:21
Operator : JC/MD
Sample : J5767-11
Misc : 6.61g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-111418-B

Quant Time: Nov 19 15:45:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

