

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111918\
 Data File : VX006092.D
 Acq On : 19 Nov 2018 15:41
 Operator : JC/MD
 Sample : J5951-04
 Misc : 5.02g/10mL/100uL/5.00ml/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FDSB-04(26.5-28.5)

Quant Time: Nov 20 00:24:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	138883	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	202542	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	166846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	89334	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	74660	46.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.52%	
35) Dibromofluoromethane	5.51	113	61945	44.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.94%	
50) Toluene-d8	8.71	98	213439	46.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.20%	
62) 4-Bromofluorobenzene	11.14	95	79690	48.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111918\
Data File : VX006092.D
Acq On : 19 Nov 2018 15:41
Operator : JC/MD
Sample : J5951-04
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FDSB-04(26.5-28.5)

Quant Time: Nov 20 00:24:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

