

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111918\
 Data File : VX006094.D
 Acq On : 19 Nov 2018 16:34
 Operator : JC/MD
 Sample : J5951-03 10X
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FDSB-04(10-12)

Quant Time: Nov 20 00:35:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	148499	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	203852	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	170867	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	92955	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	77443	45.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.72%	
35) Dibromofluoromethane	5.49	113	65237	47.05	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	94.10%	
50) Toluene-d8	8.71	98	229262	49.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.14	95	86533	52.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	2.85	84	1241	0.82	ug/l	86
31) Cyclohexane	5.58	56	8078	3.40	ug/l	# 90
39) Methylcyclohexane	7.46	83	61354	30.80	ug/l	98
73) Isopropylbenzene	11.02	105	76312	14.20	ug/l	100
78) n-propylbenzene	11.36	91	181582	28.66	ug/l	99
85) sec-Butylbenzene	11.94	105	125427	22.19	ug/l	80
89) n-Butylbenzene	12.39	91	138441	29.84	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111918\
Data File : VX006094.D
Acq On : 19 Nov 2018 16:34
Operator : JC/MD
Sample : J5951-03 10X
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FDSB-04(10-12)

Quant Time: Nov 20 00:35:48 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

