

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111918\
 Data File : VX006081.D
 Acq On : 19 Nov 2018 10:48
 Operator : JC/MD
 Sample : VX1119MBL01
 Misc : 5.0g/10mL/100uL/5.00ml/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1119MBL01

Quant Time: Nov 19 13:28:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	115890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	167752	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	152465	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	75050	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	67592	50.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.46%	
35) Dibromofluoromethane	5.51	113	55329	48.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.98%	
50) Toluene-d8	8.71	98	201283	53.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.12%	
62) 4-Bromofluorobenzene	11.14	95	68568	50.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.56%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111918\
Data File : VX006081.D
Acq On : 19 Nov 2018 10:48
Operator : JC/MD
Sample : VX1119MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1119MBL01

Quant Time: Nov 19 13:28:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX006081.D

Acq: 19 Nov 2018 10:48

Instrument :

MSVOA_X

ClientSampleId :


VX1119MBL01

Tgt Ion:168 Resp: 115890

Ion Ratio Lower Upper

168 100

99 50.7 40.7 61.1



Abundance Ion 167.90 (167.60 to 168.60): VX005825.D (-727) (-)

Ion 98.90 (98.60 to 99.60): VX005825.D (-727) (-)

5.67

40000



30000

20000

10000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 50.73 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX006081.D


Acq: 19 Nov 2018 10:48

Tgt Ion: 65 Resp: 67592

Ion Ratio Lower Upper

65 100

67 53.1 0.0 105.0



Abundance Ion 64.90 (64.60 to 65.60): VX005825.D (-795) (-)

Ion 66.90 (66.60 to 67.60): VX005825.D (-795) (-)

6.07

30000

25000

20000


15000


10000

5000

0

Time--> 6.00 6.10 6.20









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX006081.D
 Acq: 19 Nov 2018 10:48

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1119MBL01

Tgt Ion:114 Resp: 167752
 Ion Ratio Lower Upper
 114 100
 63 21.7 0.0 42.8
 88 16.2 0.0 31.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 48.49 ug/l
 RT: 5.51 min Scan# 714
 Delta R.T. -0.00 min
 Lab File: VX006081.D
 Acq: 19 Nov 2018 10:48

Tgt Ion:113 Resp: 55329
 Ion Ratio Lower Upper
 113 100
 111 101.9 82.3 123.5
 192 23.7 17.9 26.9



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 53.06 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX006081.D

Acq: 19 Nov 2018 10:48

Instrument :

MSVOA_X

ClientSampleId :


VX1119MBL01

Tgt Ion: 98 Resp: 201283

Ion Ratio Lower Upper

98 100

100 63.5 51.3 76.9



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

8.71

150000



100000

50000

0

8.60 8.70 8.80 8.90

Time-->



#62

4-Bromofluorobenzene

Concen: 50.28 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. -0.00 min

Lab File: VX006081.D

Acq: 19 Nov 2018 10:48


Tgt Ion: 95 Resp: 68568

Ion Ratio Lower Upper

95 100

174 94.0 0.0 184.4

176 91.5 0.0 177.4



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

60000


40000


20000

0

11.10 11.14 11.18 11.22

Time-->





#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 10.12 min Scan# 1470

Delta R.T. -0.00 min

Lab File: VX006081.D

Acq: 19 Nov 2018 10:48

Instrument :

MSVOA_X

ClientSampled :

VX1119MBL01


Tgt Ion:117 Resp: 152465

Ion Ratio Lower Upper

117 100

82 54.2 44.1 66.1




119 32.5 26.2 39.2



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX006081.D

Acq: 19 Nov 2018 10:48


Tgt Ion:152 Resp: 75050

Ion Ratio Lower Upper

152 100

115 55.1 39.4 118.1

150 157.1 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

