

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111919\
 Data File : VX013456.D
 Acq On : 19 Nov 2019 11:11
 Operator : JC/SP
 Sample : K5909-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Nov 19 16:22:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	731396	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1164653	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1035714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	451390	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	428828	49.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.98%	
35) Dibromofluoromethane	5.48	113	356303	48.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.56%	
50) Toluene-d8	8.71	98	1400265	50.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.70%	
62) 4-Bromofluorobenzene	11.13	95	464382	46.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.44%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111919\
Data File : VX013456.D
Acq On : 19 Nov 2019 11:11
Operator : JC/SP
Sample : K5909-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Nov 19 16:22:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

