

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111919\
 Data File : VX013461.D
 Acq On : 19 Nov 2019 13:38
 Operator : JC/SP
 Sample : K5929-23
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-GMZ6-TB1

Quant Time: Nov 19 16:30:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	731447	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1180558	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1058914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	456415	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	431992	49.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.70%	
35) Dibromofluoromethane	5.48	113	361225	48.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.56%	
50) Toluene-d8	8.71	98	1412521	50.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.20%	
62) 4-Bromofluorobenzene	11.13	95	464776	46.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111919\
Data File : VX013461.D
Acq On : 19 Nov 2019 13:38
Operator : JC/SP
Sample : K5929-23
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-GMZ6-TB1

Quant Time: Nov 19 16:30:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

