

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111920\
 Data File : VX019461.D
 Acq On : 18 Nov 2020 19:22
 Operator : JC/MD
 Sample : L4787-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB178-TB-20201113

Quant Time: Nov 19 12:04:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X111920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 16:47:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	264991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	436320	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	384331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	161371	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	183788	44.26	ug/l	0.00
Spiked Amount			Recovery	=	88.52%	
35) Dibromofluoromethane	5.46	113	131765	45.90	ug/l	0.00
Spiked Amount			Recovery	=	91.80%	
50) Toluene-d8	8.70	98	529411	47.45	ug/l	0.00
Spiked Amount			Recovery	=	94.90%	
62) 4-Bromofluorobenzene	11.12	95	181618	44.31	ug/l	0.00
Spiked Amount			Recovery	=	88.62%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111920\
Data File : VX019461.D
Acq On : 18 Nov 2020 19:22
Operator : JC/MD
Sample : L4787-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB178-TB-20201113

Quant Time: Nov 19 12:04:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X111920W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 16:47:30 2020
Response via : Initial Calibration

