

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111920\
 Data File : VX019463.D
 Acq On : 18 Nov 2020 20:09
 Operator : JC/MD
 Sample : L4787-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB178-GW-663-665

Quant Time: Nov 19 12:06:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X111920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 16:47:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	250449	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	420618	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	384969	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	165443	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	182699	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
35) Dibromofluoromethane	5.46	113	130699	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
50) Toluene-d8	8.70	98	519661	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
62) 4-Bromofluorobenzene	11.12	95	182593	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
Target Compounds						
16) Acetone	2.43	43	3734	2.460	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111920\
Data File : VX019463.D
Acq On : 18 Nov 2020 20:09
Operator : JC/MD
Sample : L4787-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB178-GW-663-665

Quant Time: Nov 19 12:06:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X111920W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 16:47:30 2020
Response via : Initial Calibration

