

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111920\
 Data File : VX019465.D
 Acq On : 18 Nov 2020 20:55
 Operator : JC/MD
 Sample : L4787-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB178-GW-718-720

Quant Time: Nov 19 12:08:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X111920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 16:47:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	253936	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	430631	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	387166	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	163540	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	185429	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
35) Dibromofluoromethane	5.46	113	132930	46.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.82%	
50) Toluene-d8	8.70	98	528940	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
62) 4-Bromofluorobenzene	11.12	95	185034	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
Target Compounds						
16) Acetone	2.42	43	14012	9.105	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111920\
Data File : VX019465.D
Acq On : 18 Nov 2020 20:55
Operator : JC/MD
Sample : L4787-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB178-GW-718-720

Quant Time: Nov 19 12:08:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X111920W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 16:47:30 2020
Response via : Initial Calibration

