


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX111920\
Data File : VX019450.D
Acq On : 18 Nov 2020 12:53
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X111920W.M
Title : SW846 8260
Last Update : Wed Nov 18 16:47:30 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1637

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	9680	PASS
75	95	30	60	52.6	26152	PASS
95	95	100	100	100.0	49680	PASS
96	95	5	9	6.9	3449	PASS
173	174	0.00	2	1.2	408	PASS
174	95	50	100	69.8	34664	PASS
175	174	5	9	7.4	2569	PASS
176	174	95	101	97.6	33840	PASS
177	176	5	9	6.6	2219	PASS