

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111924\
 Data File : VX043898.D
 Acq On : 19 Nov 2024 14:12
 Operator : JC/MD
 Sample : VX1119WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 19 23:43:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML111924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 19 23:30:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.751	114	283362	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	281248	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	115935	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	108456	47.975	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	95.940%	
7) Chloroethane-d5	1.642	69	52174	54.727	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	109.460%	
11) 1,1-Dichloroethene-d2	2.294	65	45463	50.037	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	100.080%	
21) 2-Butanone-d5	4.459	46	141678	104.289	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	104.290%	
24) Chloroform-d	5.050	84	197330	47.106	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.220%	
26) 1,2-Dichloroethane-d4	5.946	65	146053	52.520	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	105.040%	
32) Benzene-d6	5.958	84	385316	41.544	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 125	Recovery	=	83.080%	
36) 1,2-Dichloropropane-d6	7.300	67	126742	45.215	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	90.420%	
41) Toluene-d8	8.647	98	361480	44.429	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	88.860%	
43) trans-1,3-Dichloroprop...	8.946	79	53750	42.372	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	84.740%	
47) 2-Hexanone-d5	9.384	63	98039	93.676	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	93.680%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	153971	43.563	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	87.120%	
66) 1,2-Dichlorobenzene-d4	12.317	152	117094	47.582	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.160%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_X\Data\VX111924\
Data File : VX043898.D
Acq On : 19 Nov 2024 14:12
Operator : JC/MD
Sample : VX1119WBL01
Misc : 5.0mL/MSV0A_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSV0A_X
ClientSampleId :

Quant Time: Nov 19 23:43:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_X\Method\SFAMXLM111924WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 19 23:30:03 2024
Response via : Initial Calibration

