

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112018\
 Data File : VX006106.D
 Acq On : 20 Nov 2018 17:43
 Operator : JC/MD
 Sample : VSTDIC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005

Quant Time: Nov 21 07:57:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 07:49:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	317437	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	449945	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	418733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	227288	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	18065	4.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	9.82%	
35) Dibromofluoromethane	5.51	113	14167	4.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	9.80%	
50) Toluene-d8	8.71	98	53710	4.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	9.76%	
62) 4-Bromofluorobenzene	11.14	95	18817	4.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	9.10%	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.20	85	12430	5.269 ug/l	98
3) Chloromethane	1.32	50	15208	5.104 ug/l	100
4) Vinyl Chloride	1.41	62	14834	4.894 ug/l	95
5) Bromomethane	1.64	94	11174	5.317 ug/l	99
6) Chloroethane	1.71	64	8755	4.688 ug/l #	88
7) Trichlorofluoromethane	1.93	101	20476	5.164 ug/l	98
8) Diethyl Ether	2.19	74	8297	4.950 ug/l	97
9) 1,1,2-Trichlorotrifluoroet	2.38	101	12351	5.253 ug/l	99
10) Methyl Iodide	2.51	142	12487	4.236 ug/l	98
11) Tert butyl alcohol	3.07	59	19713	25.193 ug/l	95
12) 1,1-Dichloroethene	2.37	96	11549	5.039 ug/l	94
13) Acrolein	2.29	56	13282	28.367 ug/l	100
14) Allyl chloride	2.73	41	21963	4.683 ug/l	96
15) Acrylonitrile	3.15	53	42646	24.675 ug/l	100
16) Acetone	2.45	43	38556	23.622 ug/l	99
17) Carbon Disulfide	2.57	76	29664	4.793 ug/l	96
18) Methyl Acetate	2.78	43	20611	4.924 ug/l	96
19) Methyl tert-butyl Ether	3.20	73	39283	4.958 ug/l	99
20) Methylene Chloride	2.86	84	13465	5.007 ug/l	87
21) trans-1,2-Dichloroethene	3.17	96	12509	5.216 ug/l	95
22) Diisopropyl ether	3.87	45	45048	4.487 ug/l	98
23) Vinyl Acetate	3.82	43	173189	20.795 ug/l #	95
24) 1,1-Dichloroethane	3.69	63	22871	4.836 ug/l	97
25) 2-Butanone	4.70	43	70786	23.364 ug/l	99
26) 2,2-Dichloropropane	4.58	77	19161	4.916 ug/l	100
27) cis-1,2-Dichloroethene	4.60	96	16199	4.833 ug/l	97
28) Bromochloromethane	5.01	49	13173	4.779 ug/l	94
29) Tetrahydrofuran	5.15	42	43868	23.211 ug/l	98
30) Chloroform	5.21	83	28768	5.020 ug/l	95
31) Cyclohexane	5.57	56	23739	4.805 ug/l	96
32) 1,1,1-Trichloroethane	5.49	97	22559	4.875 ug/l	98
36) 1,1-Dichloropropene	5.80	75	19552	4.776 ug/l	94
37) Ethyl Acetate	4.85	43	24277	4.610 ug/l	97
38) Carbon Tetrachloride	5.78	117	18786	4.883 ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112018\
 Data File : VX006106.D
 Acq On : 20 Nov 2018 17:43
 Operator : JC/MD
 Sample : VSTDIC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005

Quant Time: Nov 21 07:57:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 07:49:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	22366	4.796	ug/l	97
40) Benzene	6.14	78	61482	5.098	ug/l	98
41) Methacrylonitrile	5.06	41	14747	5.081	ug/l	98
42) 1,2-Dichloroethane	6.20	62	21901	4.857	ug/l	98
43) Isopropyl Acetate	6.46	43	34855	4.583	ug/l	96
44) Trichloroethene	7.21	130	17264	5.151	ug/l	93
45) 1,2-Dichloropropane	7.51	63	16239	5.088	ug/l	91
46) Dibromomethane	7.66	93	10880	4.978	ug/l	99
47) Bromodichloromethane	7.90	83	19074	4.718	ug/l	97
48) Methyl methacrylate	7.77	41	16911	4.384	ug/l	95
49) 1,4-Dioxane	7.75	88	8602	93.998	ug/l	97
51) 4-Methyl-2-Pentanone	8.65	43	128547	23.576	ug/l	99
52) Toluene	8.78	92	37808	5.069	ug/l	100
53) t-1,3-Dichloropropene	9.04	75	18901	4.481	ug/l	96
54) cis-1,3-Dichloropropene	8.44	75	21620	4.636	ug/l	97
55) 1,1,2-Trichloroethane	9.21	97	15866	4.911	ug/l	98
56) Ethyl methacrylate	9.18	69	18273	4.120	ug/l	93
57) 1,3-Dichloropropane	9.37	76	26656	5.006	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.31	63	59252	23.242	ug/l	99
59) 2-Hexanone	9.49	43	100504	23.227	ug/l	96
60) Dibromochloromethane	9.58	129	14217	4.489	ug/l	98
61) 1,2-Dibromoethane	9.67	107	16321	4.904	ug/l	98
64) Tetrachloroethene	9.34	164	18614	5.402	ug/l	97
65) Chlorobenzene	10.14	112	44608	5.194	ug/l	95
66) 1,1,1,2-Tetrachloroethane	10.22	131	14150	4.799	ug/l	97
67) Ethyl Benzene	10.25	91	68731	4.887	ug/l	96
68) m/p-Xylenes	10.36	106	53159	9.707	ug/l	98
69) o-Xylene	10.70	106	25336	4.844	ug/l	100
70) Styrene	10.71	104	39278	4.605	ug/l	97
71) Bromoform	10.85	173	11477	4.348	ug/l #	98
73) Isopropylbenzene	11.02	105	67935	4.940	ug/l	98
74) N-amyl acetate	10.90	43	26119	4.184	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.27	83	25483	4.917	ug/l	98
76) 1,2,3-Trichloropropane	11.30	75	28490	6.088	ug/l	87
77) Bromobenzene	11.26	156	19422	5.041	ug/l	97
78) n-propylbenzene	11.36	91	77737	4.854	ug/l	99
79) 2-Chlorotoluene	11.42	91	49427	5.101	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	56376	4.862	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.07	75	5649	4.083	ug/l #	82
82) 4-Chlorotoluene	11.51	91	56401	4.988	ug/l	99
83) tert-Butylbenzene	11.77	119	54849	4.846	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	57203	4.837	ug/l	98
85) sec-Butylbenzene	11.94	105	69278	5.004	ug/l	99
86) p-Isopropyltoluene	12.07	119	59657	4.833	ug/l	98
87) 1,3-Dichlorobenzene	12.02	146	36308	5.111	ug/l	100
88) 1,4-Dichlorobenzene	12.10	146	37373	5.075	ug/l	93
89) n-Butylbenzene	12.39	91	51624	4.619	ug/l	98
90) Hexachloroethane	12.60	117	8917	4.527	ug/l	97
91) 1,2-Dichlorobenzene	12.39	146	36152	4.978	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	13.00	75	5367	4.463	ug/l	97


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112018\
Data File : VX006106.D
Acq On : 20 Nov 2018 17:43
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Quant Time: Nov 21 07:57:04 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 07:49:48 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.65	180	24001	4.691	ug/l	99
94) Hexachlorobutadiene	13.79	225	12571	5.042	ug/l	97
95) Naphthalene	13.83	128	64439	4.284	ug/l	99
96) 1,2,3-Trichlorobenzene	14.02	180	25189	4.830	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion:168 Resp: 317437

Ion Ratio Lower Upper




168 100

99 49.2 39.2 58.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#2

Dichlorodifluoromethane

Concen: 5.269 ug/l

RT: 1.20 min Scan# 7

Delta R.T. 0.00 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 85 Resp: 12430

Ion Ratio Lower Upper



85 100


87 30.3 15.7 47.1



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0









#3
Chloromethane
Concen: 5.104 ug/l
RT: 1.32 min Scan# 28
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 50 Resp: 15208
Ion Ratio Lower Upper
50 100
52 32.2 25.6 38.4



#4
Vinyl Chloride
Concen: 4.894 ug/l
RT: 1.41 min Scan# 42
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 62 Resp: 14834
Ion Ratio Lower Upper
62 100
64 32.9 24.2 36.4





#5

Bromomethane

Concen: 5.317 ug/l

RT: 1.64 min Scan# 79

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion: 94 Resp: 11174

Ion Ratio Lower Upper

94 100

96 94.9 76.7 115.1



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



1.64

10000

5000

0

Time-->



#6

Chloroethane

Concen: 4.688 ug/l

RT: 1.71 min Scan# 92

Delta R.T. 0.00 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 64 Resp: 8755

Ion Ratio Lower Upper

64 100

66 38.0 25.1 37.7#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

1.71


6000


4000

2000

0

Time-->









#7

Trichlorofluoromethane
Concen: 5.164 ug/l
RT: 1.93 min Scan# 127
Delta R.T. 0.01 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion: 101 Resp: 20476
Ion Ratio Lower Upper
101 100
103 63.7 52.1 78.1




#8

Diethyl Ether
Concen: 4.950 ug/l
RT: 2.19 min Scan# 170
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 74 Resp: 8297
Ion Ratio Lower Upper
74 100
45 101.3 52.2 156.6





#9

1,1,2-Trichlorotrifluoroethane

Concen: 5.253 ug/l

RT: 2.38 min Scan# 201

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005


Tgt Ion:101 Resp: 12351

Ion Ratio Lower Upper

101 100

85 45.9 36.0 54.0




151 82.8 65.8 98.6



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 4.236 ug/l

RT: 2.51 min Scan# 222

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion:142 Resp: 12487

Ion Ratio Lower Upper

142 100

127 45.3 36.0 54.0



141 16.7 11.7 17.5




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
Concen: 25.193 ug/l
RT: 3.07 min Scan# 315
Delta R.T. -0.01 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion: 59 Resp: 19713
Ion Ratio Lower Upper
59 100
57 12.8 8.7 13.1




#12

1,1-Dichloroethene
Concen: 5.039 ug/l
RT: 2.37 min Scan# 200
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 96 Resp: 11549
Ion Ratio Lower Upper
96 100
61 162.5 138.2 207.4
98 65.8 51.3 76.9





#13

Acrolein

Concen: 28.367 ug/l

RT: 2.29 min Scan# 187

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion: 56 Resp: 13282

Ion Ratio Lower Upper

56 100

55 72.1 57.8 86.6



Abundance Ion 56.00 (55.70 to 56.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

2.29

8000



6000

4000

2000

0

Time--> 2.20 2.25 2.30 2.35



#14

Allyl chloride

Concen: 4.683 ug/l

RT: 2.73 min Scan# 258

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion: 41 Resp: 21963

Ion Ratio Lower Upper

41 100

39 69.2 52.5 78.7

76 30.8 23.4 35.2



Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0

Ion 75.90 (75.60 to 76.60): VX0


15000


10000

5000

0

Time--> 2.60 2.70 2.80





#15

Acrylonitrile

Concen: 24.675 ug/l

RT: 3.15 min Scan# 327

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 53 Resp: 42646

Ion Ratio Lower Upper

53 100

52 82.7 66.4 99.6

51 37.1 29.4 44.0



#16

Acetone

Concen: 23.622 ug/l

RT: 2.45 min Scan# 212

Delta R.T. 0.00 min

Lab File: VX006106.D




Acq: 20 Nov 2018 17:43


Tgt Ion: 43 Resp: 38556

Ion Ratio Lower Upper

43 100

58 29.1 23.8 35.8









#17
Carbon Disulfide
Concen: 4.793 ug/l
RT: 2.57 min Scan# 232
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 76 Resp: 29664
Ion Ratio Lower Upper
76 100
78 8.3 7.7 11.5



#18
Methyl Acetate
Concen: 4.924 ug/l
RT: 2.78 min Scan# 266
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 43 Resp: 20611
Ion Ratio Lower Upper
43 100
74 23.4 17.3 25.9









#19
Methyl tert-butyl Ether
Concen: 4.958 ug/l
RT: 3.20 min Scan# 336
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 39283
Ion Ratio Lower Upper
73 100
57 23.2 18.2 27.4



#20
Methylene Chloride
Concen: 5.007 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.01 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 84 Resp: 13465
Ion Ratio Lower Upper
84 100
49 111.6 106.6 159.8
51 33.7 32.6 49.0
86 61.6 48.4 72.6





#21

trans-1,2-Dichloroethene

Concen: 5.216 ug/l

RT: 3.17 min Scan# 330

Delta R.T. 0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 96 Resp: 12509

Ion Ratio Lower Upper

96 100

61 145.4 122.0 183.0

98 61.9 52.7 79.1



#22

Diisopropyl ether

Concen: 4.487 ug/l

RT: 3.87 min Scan# 446

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Tgt Ion: 45 Resp: 45048




Ion Ratio Lower Upper


45 100

43 57.7 45.8 68.8

87 22.8 21.0 31.4

59 9.6 8.6 12.8





#23

Vinyl Acetate

Concen: 20.795 ug/l

RT: 3.82 min Scan# 437

Delta R.T. -0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 43 Resp: 173189

Ion Ratio Lower Upper

43 100

86 7.6 7.6 11.4#



#24

1,1-Dichloroethane

Concen: 4.836 ug/l

RT: 3.69 min Scan# 416

Delta R.T. -0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43




Tgt Ion: 63 Resp: 22871


Ion Ratio Lower Upper

63 100

98 6.2 3.3 9.8

100 3.0 2.4 7.1





#25

2-Butanone

Concen: 23.364 ug/l

RT: 4.70 min Scan# 582

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampled :


VSTDIC005

Tgt Ion: 43 Resp: 70786

Ion Ratio Lower Upper




43 100

72 23.4 19.2 28.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0



#26

2,2-Dichloropropane

Concen: 4.916 ug/l

RT: 4.58 min Scan# 562

Delta R.T. 0.00 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 77 Resp: 19161

Ion Ratio Lower Upper



77 100


97 23.5 11.9 35.5



Abundance Ion 76.90 (76.60 to 77.60): VX0

Ion 96.90 (96.60 to 97.60): VX0









#27

cis-1,2-Dichloroethene
Concen: 4.833 ug/l
RT: 4.60 min Scan# 565
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Instrument :
MSVOA_X
ClientSampled :
VSTDIC005




Tgt Ion: 96 Resp: 16199
Ion Ratio Lower Upper
96 100
61 148.2 0.0 291.6
98 69.2 0.0 128.2




#28

Bromochloromethane
Concen: 4.779 ug/l
RT: 5.01 min Scan# 633
Delta R.T. -0.01 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 49 Resp: 13173
Ion Ratio Lower Upper
49 100
129 2.6 0.0 4.4
130 82.4 61.6 92.4









#29
Tetrahydrofuran
Concen: 23.211 ug/l
RT: 5.15 min Scan# 656
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 42 Resp: 43868
Ion Ratio Lower Upper
42 100
72 40.9 33.0 49.6
71 40.9 31.0 46.4



#30
Chloroform
Concen: 5.020 ug/l
RT: 5.21 min Scan# 666
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 83 Resp: 28768
Ion Ratio Lower Upper
83 100
85 67.5 51.0 76.4





#31

Cyclohexane

Concen: 4.805 ug/l

RT: 5.57 min Scan# 724

Delta R.T. -0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 56 Resp: 23739

Ion Ratio Lower Upper

56 100

69 25.8 23.9 35.9

84 77.8 63.6 95.4



#32

1,1,1-Trichloroethane

Concen: 4.875 ug/l

RT: 5.49 min Scan# 712

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43




Tgt Ion: 97 Resp: 22559


Ion Ratio Lower Upper

97 100

99 63.9 51.6 77.4

61 47.6 36.6 55.0





#33

1,2-Dichloroethane-d4

Concen: 4.911 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion: 65 Resp: 18065

Ion Ratio Lower Upper




65 100

67 55.6 0.0 105.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion: 114 Resp: 449945

Ion Ratio Lower Upper

114 100

63 20.8 0.0 39.2



88 15.1 0.0 28.6




Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0





#35

Dibromofluoromethane

Concen: 4.905 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005


Tgt Ion: 113 Resp: 14167

Ion Ratio Lower Upper

113 100

111 105.6 81.5 122.3




192 24.9 18.6 28.0



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 4.776 ug/l

RT: 5.80 min Scan# 762

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion: 75 Resp: 19552

Ion Ratio Lower Upper

75 100

110 40.2 17.8 53.3



77 33.1 24.5 36.7




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37
Ethyl Acetate
Concen: 4.610 ug/l
RT: 4.85 min Scan# 607
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 24277
Ion Ratio Lower Upper
43 100
61 12.2 10.5 15.7
70 11.7 7.9 11.9




Abundance Ion 43.00 (42.70 to 43.70): VX0
Ion 60.90 (60.60 to 61.60): VX0
Ion 70.00 (69.70 to 70.70): VX0






#38
Carbon Tetrachloride
Concen: 4.883 ug/l
RT: 5.78 min Scan# 759
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 117 Resp: 18786
Ion Ratio Lower Upper
117 100
119 95.3 76.3 114.5
121 26.6 24.7 37.1



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V









#39
Methylcyclohexane
Concen: 4.796 ug/l
RT: 7.46 min Scan# 1034
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 22366
Ion Ratio Lower Upper
83 100
55 81.9 63.9 95.9
98 42.8 36.8 55.2




Abundance Ion 83.00 (82.70 to 83.70): VX0
Ion 55.00 (54.70 to 55.70): VX0
Ion 98.00 (97.70 to 98.70): VX0






#40
Benzene
Concen: 5.098 ug/l
RT: 6.14 min Scan# 818
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 78 Resp: 61482
Ion Ratio Lower Upper
78 100
77 24.6 18.8 28.2



Abundance Ion 78.00 (77.70 to 78.70): VX0
Ion 77.00 (76.70 to 77.70): VX0





#41

Methacrylonitrile

Concen: 5.081 ug/l

RT: 5.06 min Scan# 641

Delta R.T. 0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampled :

VSTDIC005

Tgt Ion: 41 Resp: 14747


Ion Ratio Lower Upper

41 100

39 54.4 44.1 66.1

67 64.3 54.1 81.1

52 26.3 21.7 32.5






Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0

Ion 67.00 (66.70 to 67.70): VX0

Ion 52.00 (51.70 to 52.70): VX0



#42

1,2-Dichloroethane

Concen: 4.857 ug/l

RT: 6.20 min Scan# 827

Delta R.T. 0.00 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 62 Resp: 21901

Ion Ratio Lower Upper



62 100


98 9.6 0.0 17.8



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 97.90 (97.60 to 98.60): VX0





#43

Isopropyl Acetate

Concen: 4.583 ug/l

RT: 6.46 min Scan# 870

Delta R.T. -0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 34855

Ion Ratio Lower Upper

43 100

61 20.5 15.2 22.8

87 13.8 9.6 14.4



#44

Trichloroethene

Concen: 5.151 ug/l

RT: 7.21 min Scan# 994

Delta R.T. 0.00 min

Lab File: VX006106.D




Acq: 20 Nov 2018 17:43


Tgt Ion:130 Resp: 17264

Ion Ratio Lower Upper

130 100

95 87.4 0.0 188.8





#45

1,2-Dichloropropane

Concen: 5.088 ug/l

RT: 7.51 min Scan# 1043

Delta R.T. -0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion: 63 Resp: 16239

Ion Ratio Lower Upper

63 100

65 35.7 24.8 37.2



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 64.90 (64.60 to 65.60): VX0

7.51

8000

6000



4000

2000

0

7.40 7.50 7.60

Time-->



#46

Dibromomethane

Concen: 4.978 ug/l

RT: 7.66 min Scan# 1068

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion: 93 Resp: 10880

Ion Ratio Lower Upper

93 100

95 81.4 66.3 99.5

174 113.6 91.0 136.6



Abundance Ion 92.80 (92.50 to 93.50): VX0

Ion 94.80 (94.50 to 95.50): VX0

Ion 173.70 (173.40 to 174.40): V

8000

6000


4000


2000

0

7.60 7.65 7.70 7.75

Time-->









#47
 Bromodichloromethane
 Concen: 4.718 ug/l
 RT: 7.90 min Scan# 1106
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43




Instrument :
 MSVOA_X
 ClientSampled :
 VSTDIC005


Tgt Ion: 83 Resp: 19074
 Ion Ratio Lower Upper
 83 100
 85 65.7 50.4 75.6
 127 8.8 7.0 10.4



#48
 Methyl methacrylate
 Concen: 4.384 ug/l
 RT: 7.77 min Scan# 1086
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43

Tgt Ion: 41 Resp: 16911
 Ion Ratio Lower Upper
 41 100
 69 82.2 62.8 94.2
 39 59.1 43.8 65.6





#49

1,4-Dioxane

Concen: 93.998 ug/l

RT: 7.75 min Scan# 1082

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005


Tgt Ion: 88 Resp: 8602

Ion Ratio Lower Upper

88 100

43 38.7 26.7 40.1




58 73.0 58.5 87.7



Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 4.875 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 98 Resp: 53710

Ion Ratio Lower Upper



98 100


100 64.3 51.0 76.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0









#51
 4-Methyl-2-Pentanone
 Concen: 23.576 ug/l
 RT: 8.65 min Scan# 1229
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 43 Resp: 128547
 Ion Ratio Lower Upper
 43 100
 58 36.9 29.9 44.9



#52
 Toluene
 Concen: 5.069 ug/l
 RT: 8.78 min Scan# 1251
 Delta R.T. -0.01 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43

Tgt Ion: 92 Resp: 37808
 Ion Ratio Lower Upper
 92 100
 91 178.3 142.5 213.7





#53

t-1,3-Dichloropropene

Concen: 4.481 ug/l

RT: 9.04 min Scan# 1294

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion: 75 Resp: 18901

Ion Ratio Lower Upper

75 100

77 30.0 25.7 38.5



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 76.90 (76.60 to 77.60): VX0

9.04

15000



10000

5000

0

Time-->

9.00 9.05 9.10



#54

cis-1,3-Dichloropropene

Concen: 4.636 ug/l

RT: 8.44 min Scan# 1195

Delta R.T. 0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion: 75 Resp: 21620

Ion Ratio Lower Upper

75 100

77 33.3 24.7 37.1

39 48.9 40.3 60.5



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 76.90 (76.60 to 77.60): VX0

Ion 39.00 (38.70 to 39.70): VX0

8.44

15000


10000


5000

0

Time-->

8.35 8.40 8.45 8.50









#55
 1,1,2-Trichloroethane
 Concen: 4.911 ug/l
 RT: 9.21 min Scan# 1322
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005

Tgt Ion:	97	Resp:	15866
Ion	Ratio	Lower	Upper
97	100		
83	85.5	70.7	106.1
85	58.3	45.7	68.5
99	64.8	51.4	77.2




Abundance Ion 96.90 (96.60 to 97.60): VX0
 Ion 82.90 (82.60 to 83.60): VX0
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 98.90 (98.60 to 99.60): VX0






#56
 Ethyl methacrylate
 Concen: 4.120 ug/l
 RT: 9.18 min Scan# 1317
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43

Tgt Ion:	69	Resp:	18273
Ion	Ratio	Lower	Upper
69	100		
41	75.4	56.2	84.4
39	41.0	29.4	44.0



Abundance Ion 69.00 (68.70 to 69.70): VX0
 Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0





#57

1,3-Dichloropropane

Concen: 5.006 ug/l

RT: 9.37 min Scan# 1348

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion: 76 Resp: 26656

Ion Ratio Lower Upper




76 100

78 31.4 25.7 38.5



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0



#58

2-Chloroethyl Vinyl ether

Concen: 23.242 ug/l

RT: 8.31 min Scan# 1174

Delta R.T. -0.01 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 63 Resp: 59252

Ion Ratio Lower Upper



63 100


106 26.8 21.7 32.5



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 105.90 (105.60 to 106.60): V









#59
2-Hexanone
Concen: 23.227 ug/l
RT: 9.49 min Scan# 1368
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 100504
Ion Ratio Lower Upper
43 100
58 49.8 26.3 78.9



#60
Dibromochloromethane
Concen: 4.489 ug/l
RT: 9.58 min Scan# 1382
Delta R.T. -0.01 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 129 Resp: 14217
Ion Ratio Lower Upper
129 100
127 80.3 39.1 117.5





#61

1,2-Dibromoethane

Concen: 4.904 ug/l

RT: 9.67 min Scan# 1397

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion: 107 Resp: 16321

Ion Ratio Lower Upper




107 100

109 96.3 75.7 113.5



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



#62

4-Bromofluorobenzene

Concen: 4.552 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion: 95 Resp: 18817

Ion Ratio Lower Upper

95 100

174 92.8 0.0 184.4



176 88.6 0.0 179.0




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:117 Resp: 418733
Ion Ratio Lower Upper
117 100
82 53.4 43.0 64.4
119 32.3 26.1 39.1



#64
Tetrachloroethene
Concen: 5.402 ug/l
RT: 9.34 min Scan# 1342
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion:164 Resp: 18614
Ion Ratio Lower Upper
164 100
166 128.9 101.5 152.3
129 95.2 73.4 110.0
131 91.6 70.2 105.4





#65

Chlorobenzene

Concen: 5.194 ug/l

RT: 10.14 min Scan# 1474

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion:112 Resp: 44608

Ion Ratio Lower Upper




112 100

114 29.3 25.6 38.4



Abundance Ion 111.90 (111.60 to 112.60): V

Ion 113.90 (113.60 to 114.60): V



#66

1,1,1,2-Tetrachloroethane

Concen: 4.799 ug/l

RT: 10.22 min Scan# 1487

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion:131 Resp: 14150

Ion Ratio Lower Upper

131 100

133 98.0 48.0 144.2



119 68.7 32.6 97.8




Abundance Ion 130.80 (130.50 to 131.50): V

Ion 132.80 (132.50 to 133.50): V

Ion 118.80 (118.50 to 119.50): V





#67

Ethyl Benzene

Concen: 4.887 ug/l

RT: 10.25 min Scan# 1492

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 91 Resp: 68731

Ion Ratio Lower Upper

91 100

106 32.9 24.4 36.6



#68

m/p-Xylenes

Concen: 9.707 ug/l

RT: 10.36 min Scan# 1510

Delta R.T. 0.00 min

Lab File: VX006106.D




Acq: 20 Nov 2018 17:43


Tgt Ion: 106 Resp: 53159

Ion Ratio Lower Upper

106 100

91 198.8 162.0 243.0









#69
o-Xylene
Concen: 4.844 ug/l
RT: 10.70 min Scan# 1566
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:106 Resp: 25336
Ion Ratio Lower Upper
106 100
91 214.5 107.4 322.2



#70
Styrene
Concen: 4.605 ug/l
RT: 10.71 min Scan# 1568
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion:104 Resp: 39278
Ion Ratio Lower Upper
104 100
78 50.6 39.4 59.2
103 58.8 44.6 67.0





#71

Bromoform

Concen: 4.348 ug/l

RT: 10.85 min Scan# 1591

Delta R.T. -0.01 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampled :

VSTDIC005


Tgt Ion:173 Resp: 11477

Ion Ratio Lower Upper

173 100

175 48.0 24.9 74.6




254 0.3 0.2 0.2#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion:152 Resp: 227288

Ion Ratio Lower Upper

152 100

115 57.7 39.0 116.9



150 155.7 0.0 346.6




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V









#73
Isopropylbenzene
Concen: 4.940 ug/l
RT: 11.02 min Scan# 1618
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 105 Resp: 67935
Ion Ratio Lower Upper
105 100
120 27.6 13.3 39.8



#74
N-ethyl acetate
Concen: 4.184 ug/l
RT: 10.90 min Scan# 1599
Delta R.T. 0.01 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 43 Resp: 26119
Ion Ratio Lower Upper
43 100
70 40.6 31.8 47.6
55 27.9 19.5 29.3
61 22.0 18.6 27.8





#75

1,1,2,2-Tetrachloroethane

Concen: 4.917 ug/l

RT: 11.27 min Scan# 1659

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005


Tgt Ion: 83 Resp: 25483

Ion Ratio Lower Upper

83 100

131 11.4 5.2 15.6




85 66.5 32.6 97.8



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): VX0

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 6.088 ug/l

RT: 11.30 min Scan# 1664

Delta R.T. 0.00 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 75 Resp: 28490

Ion Ratio Lower Upper



75 100


77 30.9 19.6 58.7



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 5.041 ug/l

RT: 11.26 min Scan# 1657

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampled :

VSTDIC005


Tgt Ion:156 Resp: 19422

Ion Ratio Lower Upper

156 100

77 143.3 74.3 222.8




158 97.2 49.0 147.0



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 4.854 ug/l

RT: 11.36 min Scan# 1674

Delta R.T. 0.00 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 91 Resp: 77737

Ion Ratio Lower Upper



91 100


120 23.1 11.8 35.3



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 5.101 ug/l

RT: 11.42 min Scan# 1684

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion: 91 Resp: 49427

Ion Ratio Lower Upper




91 100

126 34.8 17.6 52.9



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V



#80

1,3,5-Trimethylbenzene

Concen: 4.862 ug/l

RT: 11.51 min Scan# 1698

Delta R.T. 0.00 min

Lab File: VX006106.D


Acq: 20 Nov 2018 17:43

Tgt Ion: 105 Resp: 56376

Ion Ratio Lower Upper



105 100


120 49.0 24.3 73.0



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 4.083 ug/l

RT: 11.07 min Scan# 1627

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 75 Resp: 5649

Ion Ratio Lower Upper

75 100

53 132.8 85.3 127.9#

89 47.0 37.2 55.8



#82

4-Chlorotoluene

Concen: 4.988 ug/l

RT: 11.51 min Scan# 1699

Delta R.T. 0.00 min

Lab File: VX006106.D




Acq: 20 Nov 2018 17:43


Tgt Ion: 91 Resp: 56401

Ion Ratio Lower Upper

91 100

126 30.9 15.3 45.9









#83
 tert-Butylbenzene
 Concen: 4.846 ug/l
 RT: 11.77 min Scan# 1741
 Delta R.T. -0.01 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43




Instrument :
 MSVOA_X
 ClientSampled :
 VSTDIC005


Tgt Ion:119 Resp: 54849
 Ion Ratio Lower Upper
 119 100
 91 58.6 28.2 84.6
 134 24.5 11.5 34.4



#84
 1,2,4-Trimethylbenzene
 Concen: 4.837 ug/l
 RT: 11.81 min Scan# 1748
 Delta R.T. 0.01 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43

Tgt Ion:105 Resp: 57203
 Ion Ratio Lower Upper
 105 100
 120 44.2 22.7 68.0





#85

sec-Butylbenzene

Concen: 5.004 ug/l

RT: 11.94 min Scan# 1770

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43

Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Tgt Ion:105 Resp: 69278

Ion Ratio Lower Upper




105 100

134 20.1 10.4 31.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 4.833 ug/l

RT: 12.07 min Scan# 1790

Delta R.T. 0.00 min

Lab File: VX006106.D

Acq: 20 Nov 2018 17:43


Tgt Ion:119 Resp: 59657

Ion Ratio Lower Upper

119 100

134 26.6 13.1 39.3



91 24.2 11.3 33.9




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VX0









#87
 1,3-Dichlorobenzene
 Concen: 5.111 ug/l
 RT: 12.02 min Scan# 1783
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:146 Resp: 36308
 Ion Ratio Lower Upper
 146 100
 111 37.6 18.8 56.3
 148 64.2 32.2 96.6



#88
 1,4-Dichlorobenzene
 Concen: 5.075 ug/l
 RT: 12.10 min Scan# 1795
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43

Tgt Ion:146 Resp: 37373
 Ion Ratio Lower Upper
 146 100
 111 42.6 18.3 54.9
 148 68.2 32.2 96.6









#89
n-Butylbenzene
Concen: 4.619 ug/l
RT: 12.39 min Scan# 1843
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 51624
Ion Ratio Lower Upper
91 100
92 51.2 26.6 79.7
134 26.2 13.1 39.3



#90
Hexachloroethane
Concen: 4.527 ug/l
RT: 12.60 min Scan# 1877
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion: 117 Resp: 8917
Ion Ratio Lower Upper
117 100
201 88.9 46.1 138.2









#91
 1,2-Dichlorobenzene
 Concen: 4.978 ug/l
 RT: 12.39 min Scan# 1843
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:146 Resp: 36152
 Ion Ratio Lower Upper
 146 100
 111 40.1 19.6 58.8
 148 65.4 32.0 96.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 4.463 ug/l
 RT: 13.00 min Scan# 1943
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43

Tgt Ion: 75 Resp: 5367
 Ion Ratio Lower Upper
 75 100
 155 87.1 46.1 138.3
 157 118.2 59.8 179.3









#93
 1,2,4-Trichlorobenzene
 Concen: 4.691 ug/l
 RT: 13.65 min Scan# 2049
 Delta R.T. 0.00 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:180 Resp: 24001
 Ion Ratio Lower Upper
 180 100
 182 95.8 48.2 144.6
 145 28.6 14.1 42.2



#94
 Hexachlorobutadiene
 Concen: 5.042 ug/l
 RT: 13.79 min Scan# 2072
 Delta R.T. 0.01 min
 Lab File: VX006106.D
 Acq: 20 Nov 2018 17:43

Tgt Ion:225 Resp: 12571
 Ion Ratio Lower Upper
 225 100
 223 63.8 31.2 93.6
 227 67.5 32.4 97.2









#95
Naphthalene
Concen: 4.284 ug/l
RT: 13.83 min Scan# 2080
Delta R.T. 0.00 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:128 Resp: 64439
Ion Ratio Lower Upper
128 100
127 13.3 10.6 15.8
129 11.4 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 4.830 ug/l
RT: 14.02 min Scan# 2111
Delta R.T. 0.01 min
Lab File: VX006106.D
Acq: 20 Nov 2018 17:43

Tgt Ion:180 Resp: 25189
Ion Ratio Lower Upper
180 100
182 96.1 47.5 142.5
145 28.7 15.0 45.0

