


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112018\
Data File : VX006104.D
Acq On : 20 Nov 2018 14:48
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Title : SW846 8260
Last Update : Wed Nov 21 08:15:54 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.7	3072	PASS
75	95	30	60	58.4	7266	PASS
95	95	100	100	100.0	12442	PASS
96	95	5	9	6.7	831	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	93.5	11630	PASS
175	174	5	9	8.1	947	PASS
176	174	95	101	99.4	11555	PASS
177	176	5	9	7.2	832	PASS