

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112019\
 Data File : VX013472.D
 Acq On : 20 Nov 2019 12:00
 Operator : JC/SP
 Sample : K5949-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-191119

Quant Time: Nov 21 01:17:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	708382	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1150427	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1015112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	462908	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	414417	49.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.76%	
35) Dibromofluoromethane	5.48	113	348648	48.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.64%	
50) Toluene-d8	8.71	98	1375120	50.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.10%	
62) 4-Bromofluorobenzene	11.14	95	455785	46.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.86%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112019\
Data File : VX013472.D
Acq On : 20 Nov 2019 12:00
Operator : JC/SP
Sample : K5949-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-191119

Quant Time: Nov 21 01:17:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

