

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112019\
 Data File : VX013485.D
 Acq On : 20 Nov 2019 18:18
 Operator : JC/SP
 Sample : K5957-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-SOIL-111919

Quant Time: Nov 21 04:30:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	721397	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1155821	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1031828	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	450393	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	420687	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
35) Dibromofluoromethane	5.49	113	351586	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
50) Toluene-d8	8.71	98	1367781	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.14	95	458518	46.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.98%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	11.81	105	45191	1.811	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112019\
Data File : VX013485.D
Acq On : 20 Nov 2019 18:18
Operator : JC/SP
Sample : K5957-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-SOIL-111919

Quant Time: Nov 21 04:30:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

