

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112020\
 Data File : VX019478.D
 Acq On : 19 Nov 2020 20:04
 Operator : JC/MD
 Sample : L4827-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 201118001-05-TRIP-BLANK

Quant Time: Nov 20 02:11:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X112020W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 20 01:06:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.98	128	52280	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.83	114	285187	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	259483	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.03	65	114973	30.60	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	102.00%		
60) 4-Bromofluorobenzene	11.12	95	115685	27.34	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	91.13%		
63) Toluene-d8	8.70	98	341043	29.49	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	98.30%		

Target Compounds

					Ovalue
15) Acetone	2.43	58	1278	2.600 ug/l #	64
16) Carbon Disulfide	2.54	76	2992	0.391 ug/l #	92
18) Methylene Chloride	2.83	84	3024	0.864 ug/l	95
25) Chloroform	5.17	83	6739	1.129 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112020\
Data File : VX019478.D
Acq On : 19 Nov 2020 20:04
Operator : JC/MD
Sample : L4827-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
201118001-05-TRIP-BLANK

Quant Time: Nov 20 02:11:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X112020W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 20 01:06:02 2020
Response via : Initial Calibration

