

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
 Data File : VX038952.D
 Acq On : 20 Nov 2023 12:47
 Operator : JC/MD
 Sample : 05428-02 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20231113

Quant Time: Nov 21 00:18:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	120812	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	246203	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	252760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	126247	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95607	49.112	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.220%		
35) Dibromofluoromethane	5.385	113	78417	44.101	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.200%		
50) Toluene-d8	8.647	98	306884	45.109	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	90.220%#		
62) 4-Bromofluorobenzene	11.079	95	130872	48.055	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.100%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	2274	1.525	ug/l	94
12) 1,1-Dichloroethene	2.325	96	1433	0.952	ug/l	85
20) Methylene Chloride	2.788	84	5297	2.836	ug/l	93
44) Trichloroethene	7.123	130	128659	70.124	ug/l	93
64) Tetrachloroethene	9.275	164	1256	0.799	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
Data File : VX038952.D
Acq On : 20 Nov 2023 12:47
Operator : JC/MD
Sample : 05428-02 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20231113

Quant Time: Nov 21 00:18:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

