

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
 Data File : VX038954.D
 Acq On : 20 Nov 2023 13:33
 Operator : JC/MD
 Sample : 05404-31DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D2-20231110DL

Quant Time: Nov 21 00:19:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	112533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228809	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89333	49.266	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.540%		
35) Dibromofluoromethane	5.385	113	73627	44.555	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.100%		
50) Toluene-d8	8.646	98	289260	45.750	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	91.500%#		
62) 4-Bromofluorobenzene	11.079	95	126382	49.934	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	99.860%		
Target Compounds						
20) Methylene Chloride	2.788	84	3653	2.099	ug/l	90
44) Trichloroethene	7.122	130	51326	30.101	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
Data File : VX038954.D
Acq On : 20 Nov 2023 13:33
Operator : JC/MD
Sample : 05404-31DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D2-20231110DL

Quant Time: Nov 21 00:19:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

