

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
 Data File : VX038959.D
 Acq On : 20 Nov 2023 15:28
 Operator : JC/MD
 Sample : 05431-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-20231113

Quant Time: Nov 21 00:21:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220594	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87460	49.906	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.820%
35) Dibromofluoromethane	5.385	113	70931	44.522	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.040%
50) Toluene-d8	8.647	98	282543	46.352	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.700%
62) 4-Bromofluorobenzene	11.079	95	120728	49.476	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
Data File : VX038959.D
Acq On : 20 Nov 2023 15:28
Operator : JC/MD
Sample : 05431-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-20231113

Quant Time: Nov 21 00:21:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

